

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070822\
 Data File : VN073404.D
 Acq On : 08 Jul 2022 19:46
 Operator : JC\MD
 Sample : N3642-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-2

Quant Time: Jul 09 05:30:15 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	289661	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	495055	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	449624	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	224881	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	190399	45.533	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	91.060%
35) Dibromofluoromethane	7.951	113	159524	48.433	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	96.860%
50) Toluene-d8	10.380	98	518602	44.333	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	88.660%
62) 4-Bromofluorobenzene	12.669	95	227530	50.613	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	101.220%
Target Compounds						
					Qvalue	
8) Diethyl Ether	3.775	74	1356	0.552	ug/l	70
27) cis-1,2-Dichloroethene	7.251	96	1577	0.365	ug/l	81
65) Chlorobenzene	11.710	112	5165	0.508	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070822\
Data File : VN073404.D
Acq On : 08 Jul 2022 19:46
Operator : JC\MD
Sample : N3642-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-2

Quant Time: Jul 09 05:30:15 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

