

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070824\
 Data File : VN082815.D
 Acq On : 08 Jul 2024 11:34
 Operator : JC\MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC020

Quant Time: Jul 08 13:22:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:14:54 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	159007	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	267615	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	241012	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	110512	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	44843	20.118	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	40.240%#
35) Dibromofluoromethane	8.165	113	34532	20.109	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	40.220%#
50) Toluene-d8	10.571	98	126378	19.926	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	39.860%#
62) 4-Bromofluorobenzene	12.847	95	45541	19.989	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	39.980%#
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	2.124	85	35772	17.377	ug/l	99
3) Chloromethane	2.360	50	34794	17.822	ug/l	93
4) Vinyl Chloride	2.512	62	39204	19.121	ug/l	98
5) Bromomethane	2.942	94	29103	19.687	ug/l	98
6) Chloroethane	3.112	64	27048	19.277	ug/l	94
7) Trichlorofluoromethane	3.489	101	54177	18.371	ug/l	96
8) Diethyl Ether	3.965	74	19474	19.295	ug/l	79
9) 1,1,2-Trichlorotrifluo...	4.365	101	31417	18.823	ug/l	98
10) Methyl Iodide	4.583	142	30653	17.913	ug/l	93
11) Tert butyl alcohol	5.524	59	36701	97.896	ug/l	96
12) 1,1-Dichloroethene	4.330	96	30062	18.520	ug/l #	80
13) Acrolein	4.177	56	25388	93.798	ug/l	97
14) Allyl chloride	5.024	41	49734	19.651	ug/l	91
15) Acrylonitrile	5.718	53	84251	97.787	ug/l	99
16) Acetone	4.430	43	69999	89.198	ug/l	94
17) Carbon Disulfide	4.712	76	84689	17.460	ug/l	98
18) Methyl Acetate	5.030	43	36819	17.731	ug/l #	88
19) Methyl tert-butyl Ether	5.795	73	110099	19.663	ug/l	95
20) Methylene Chloride	5.277	84	35314	18.633	ug/l #	84
21) trans-1,2-Dichloroethene	5.783	96	34060	19.366	ug/l	86
22) Diisopropyl ether	6.677	45	106056	20.030	ug/l	97
23) Vinyl Acetate	6.606	43	425025	101.080	ug/l #	93
24) 1,1-Dichloroethane	6.571	63	63543	19.785	ug/l	97
25) 2-Butanone	7.483	43	117674	99.308	ug/l #	90
26) 2,2-Dichloropropane	7.489	77	60874	20.385	ug/l	98
27) cis-1,2-Dichloroethene	7.489	96	39763	19.311	ug/l	92
28) Bromochloromethane	7.812	49	26971	19.575	ug/l #	78
29) Tetrahydrofuran	7.841	42	74830	100.310	ug/l #	86
30) Chloroform	7.965	83	65971	19.362	ug/l	99
31) Cyclohexane	8.253	56	56951	19.543	ug/l	91
32) 1,1,1-Trichloroethane	8.165	97	62884	19.878	ug/l	96
36) 1,1-Dichloropropene	8.371	75	47023	19.861	ug/l	98
37) Ethyl Acetate	7.559	43	44783	18.579	ug/l #	94
38) Carbon Tetrachloride	8.365	117	55714	19.544	ug/l	98
39) Methylcyclohexane	9.600	83	49844	19.490	ug/l	93
40) Benzene	8.606	78	141624	19.446	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070824\
 Data File : VN082815.D
 Acq On : 08 Jul 2024 11:34
 Operator : JC\MD
 Sample : VSTDICC020
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC020

Quant Time: Jul 08 13:22:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:14:54 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.783	41	26613	20.494	ug/l	96
42) 1,2-Dichloroethane	8.677	62	52120	19.879	ug/l	99
43) Isopropyl Acetate	8.688	43	102312	20.889	ug/l	98
44) Trichloroethene	9.353	130	37261	19.970	ug/l	93
45) 1,2-Dichloropropane	9.624	63	34621	20.362	ug/l	98
46) Dibromomethane	9.712	93	25968	19.605	ug/l	97
47) Bromodichloromethane	9.888	83	53763	19.644	ug/l	99
48) Methyl methacrylate	9.683	41	37663	19.927	ug/l #	88
49) 1,4-Dioxane	9.700	88	16625	405.531	ug/l	91
51) 4-Methyl-2-Pentanone	10.447	43	240308	100.562	ug/l	93
52) Toluene	10.630	92	91596	20.134	ug/l	98
53) t-1,3-Dichloropropene	10.835	75	56950	20.199	ug/l	98
54) cis-1,3-Dichloropropene	10.312	75	58926	19.745	ug/l	89
55) 1,1,2-Trichloroethane	11.018	97	34086	20.375	ug/l	97
56) Ethyl methacrylate	10.877	69	53736	20.647	ug/l	88
57) 1,3-Dichloropropane	11.165	76	57819	19.525	ug/l	97
58) 2-Chloroethyl Vinyl ether	10.159	63	134913	101.020	ug/l	94
59) 2-Hexanone	11.200	43	184214	103.685	ug/l	92
60) Dibromochloromethane	11.359	129	42676	20.029	ug/l	99
61) 1,2-Dibromoethane	11.471	107	36253	20.406	ug/l	100
64) Tetrachloroethene	11.106	164	35092	19.516	ug/l	98
65) Chlorobenzene	11.894	112	102037	19.757	ug/l	97
66) 1,1,1,2-Tetrachloroethane	11.959	131	36936	20.107	ug/l	99
67) Ethyl Benzene	11.965	91	171823	19.795	ug/l	96
68) m/p-Xylenes	12.077	106	133538	40.750	ug/l	97
69) o-Xylene	12.400	106	65258	20.783	ug/l	95
70) Styrene	12.412	104	106102	20.451	ug/l	100
71) Bromoform	12.582	173	29766	20.604	ug/l #	99
73) Isopropylbenzene	12.694	105	166182	20.694	ug/l	98
74) N-amyl acetate	12.494	43	69100	20.803	ug/l	93
75) 1,1,2,2-Tetrachloroethane	12.941	83	48149	20.020	ug/l	99
76) 1,2,3-Trichloropropane	12.994	75	67963	31.090	ug/l	41
77) Bromobenzene	12.982	156	41591	20.076	ug/l	93
78) n-propylbenzene	13.035	91	188334	20.398	ug/l	99
79) 2-Chlorotoluene	13.129	91	117405	19.957	ug/l	98
80) 1,3,5-Trimethylbenzene	13.176	105	135106	20.814	ug/l	99
81) trans-1,4-Dichloro-2-b...	12.735	75	19473	20.101	ug/l	97
82) 4-Chlorotoluene	13.224	91	116588	20.022	ug/l	97
83) tert-Butylbenzene	13.441	119	117887	20.486	ug/l	95
84) 1,2,4-Trimethylbenzene	13.482	105	138273	20.900	ug/l	100
85) sec-Butylbenzene	13.618	105	160246	20.607	ug/l	99
86) p-Isopropyltoluene	13.729	119	132739	20.459	ug/l	99
87) 1,3-Dichlorobenzene	13.735	146	73937	19.756	ug/l	100
88) 1,4-Dichlorobenzene	13.812	146	74739	19.901	ug/l	98
89) n-Butylbenzene	14.059	91	115371	20.756	ug/l	100
90) Hexachloroethane	14.335	117	27232	19.836	ug/l	99
91) 1,2-Dichlorobenzene	14.106	146	72440	19.936	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.723	75	11044	19.849	ug/l	93
93) 1,2,4-Trichlorobenzene	15.394	180	39420	20.140	ug/l	99
94) Hexachlorobutadiene	15.500	225	17310	19.950	ug/l	98
95) Naphthalene	15.641	128	134862	20.117	ug/l	99
96) 1,2,3-Trichlorobenzene	15.841	180	39984	19.868	ug/l	98


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070824\
Data File : VN082815.D
Acq On : 08 Jul 2024 11:34
Operator : JC\MD
Sample : VSTDICC020
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Quant Time: Jul 08 13:22:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:14:54 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------





(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.230 min Scan# 10
Delta R.T. 0.001 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020

Tgt Ion:168 Resp: 159007
Ion Ratio Lower Upper
168 100
99 55.7 48.2 72.4




#2
Dichlorodifluoromethane
Concen: 17.377 ug/l
RT: 2.124 min Scan# 29
Delta R.T. 0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion: 85 Resp: 35772
Ion Ratio Lower Upper
85 100
87 30.6 15.7 47.0



Abundance Scan 69 (2.359 min): VN082598.D\data.ms (-62)



#3

Chloromethane

Concen: 17.822 ug/l

RT: 2.360 min Scan# 69

Delta R.T. 0.001 min

Lab File: VN082815.D

Acq: 08 Jul 2024 11:34

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020


Tgt Ion: 50 Resp: 34794

Ion Ratio Lower Upper


50 100

52 35.3 25.0 37.4


Abundance Scan 69 (2.360 min): VN082815.D\data.ms




Abundance



Abundance Scan 69 (2.360 min): VN082815.D\data.ms (-18)



Abundance Scan 95 (2.512 min): VN082598.D\data.ms (-87)



#4

Vinyl Chloride

Concen: 19.121 ug/l

RT: 2.512 min Scan# 95

Delta R.T. 0.000 min

Lab File: VN082815.D

Acq: 08 Jul 2024 11:34


Tgt Ion: 62 Resp: 39204

Ion Ratio Lower Upper


62 100

64 30.0 25.0 37.6


Abundance Scan 95 (2.512 min): VN082815.D\data.ms




Abundance



Abundance Scan 95 (2.512 min): VN082815.D\data.ms (-44)









#5
Bromomethane
Concen: 19.687 ug/l
RT: 2.942 min Scan# 1
Delta R.T. -0.005 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 94 Resp: 29103
Ion Ratio Lower Upper
94 100
96 95.6 78.0 117.0



#6
Chloroethane
Concen: 19.277 ug/l
RT: 3.112 min Scan# 197
Delta R.T. 0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion: 64 Resp: 27048
Ion Ratio Lower Upper
64 100
66 29.7 26.6 40.0









#7
Trichlorofluoromethane
Concen: 18.371 ug/l
RT: 3.489 min Scan# 262
Delta R.T. -0.005 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion:101 Resp: 54177
Ion Ratio Lower Upper
101 100
103 62.5 52.7 79.1



#8
Diethyl Ether
Concen: 19.295 ug/l
RT: 3.965 min Scan# 342
Delta R.T. 0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion: 74 Resp: 19474
Ion Ratio Lower Upper
74 100
45 88.9 55.5 166.3









#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.823 ug/l
RT: 4.365 min Scan# 411
Delta R.T. -0.012 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020

Tgt Ion:101 Resp: 31417




Ion	Ratio	Lower	Upper
101	100		
85	46.3	38.3	57.5
151	75.1	59.3	88.9




#10
Methyl Iodide
Concen: 17.913 ug/l
RT: 4.583 min Scan# 447
Delta R.T. -0.005 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion:142 Resp: 30653

Ion	Ratio	Lower	Upper
142	100		
127	41.9	37.5	56.3
141	14.2	13.1	19.7









#11
Tert butyl alcohol
Concen: 97.896 ug/l
RT: 5.524 min Scan# 607
Delta R.T. 0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 59 Resp: 36701
Ion Ratio Lower Upper
59 100
57 9.4 8.9 13.3



#12
1,1-Dichloroethene
Concen: 18.520 ug/l
RT: 4.330 min Scan# 404
Delta R.T. -0.011 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion: 96 Resp: 30062
Ion Ratio Lower Upper
96 100
61 148.3 149.7 224.5#
98 61.3 50.1 75.1









#13
Acrolein
Concen: 93.798 ug/l
RT: 4.177 min Scan# 31
Delta R.T. 0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 56 Resp: 25388
Ion Ratio Lower Upper
56 100
55 73.1 56.4 84.6



#14
Allyl chloride
Concen: 19.651 ug/l
RT: 5.024 min Scan# 522
Delta R.T. 0.006 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion: 41 Resp: 49734
Ion Ratio Lower Upper
41 100
39 76.6 67.8 101.6
76 37.6 25.8 38.8









#15
Acrylonitrile
Concen: 97.787 ug/l
RT: 5.718 min Scan# 640
Delta R.T. 0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion:	53	Resp:	84251
Ion	Ratio	Lower	Upper
53	100		
52	82.7	66.6	100.0
51	35.8	30.1	45.1



#16
Acetone
Concen: 89.198 ug/l
RT: 4.430 min Scan# 421
Delta R.T. -0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion:	43	Resp:	69999
Ion	Ratio	Lower	Upper
43	100		
58	30.3	21.8	32.6









#17
Carbon Disulfide
Concen: 17.460 ug/l
RT: 4.712 min Scan# 469
Delta R.T. 0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 76 Resp: 84689
Ion Ratio Lower Upper
76 100
78 9.6 7.2 10.8



#18
Methyl Acetate
Concen: 17.731 ug/l
RT: 5.030 min Scan# 523
Delta R.T. -0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion: 43 Resp: 36819
Ion Ratio Lower Upper
43 100
74 25.9 16.3 24.5#









#19
Methyl tert-butyl Ether
Concen: 19.663 ug/l
RT: 5.795 min Scan# 61
Delta R.T. -0.005 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 73 Resp: 110099
Ion Ratio Lower Upper
73 100
57 19.8 17.9 26.9



#20
Methylene Chloride
Concen: 18.633 ug/l
RT: 5.277 min Scan# 565
Delta R.T. -0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion: 84 Resp: 35314
Ion Ratio Lower Upper
84 100
49 121.7 119.6 179.4
51 34.3 34.8 52.2#
86 63.5 52.9 79.3









#21
trans-1,2-Dichloroethene
Concen: 19.366 ug/l
RT: 5.783 min Scan# 61
Delta R.T. -0.005 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 96 Resp: 34060
Ion Ratio Lower Upper
96 100
61 129.6 123.4 185.2
98 60.9 49.3 73.9



#22
Diisopropyl ether
Concen: 20.030 ug/l
RT: 6.677 min Scan# 803
Delta R.T. -0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion: 45 Resp: 106056
Ion Ratio Lower Upper
45 100
43 55.0 44.0 66.0
87 29.3 19.7 29.5
59 12.0 8.5 12.7









#23
 Vinyl Acetate
 Concen: 101.080 ug/l
 RT: 6.606 min Scan# 791
 Delta R.T. 0.000 min
 Lab File: VN082815.D
 Acq: 08 Jul 2024 11:34




Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDICC020


Tgt Ion: 43 Resp: 425025
 Ion Ratio Lower Upper
 43 100
 86 11.7 7.3 10.9#



#24
 1,1-Dichloroethane
 Concen: 19.785 ug/l
 RT: 6.571 min Scan# 785
 Delta R.T. -0.000 min
 Lab File: VN082815.D
 Acq: 08 Jul 2024 11:34

Tgt Ion: 63 Resp: 63543
 Ion Ratio Lower Upper
 63 100
 98 7.9 3.3 9.9
 100 4.6 2.0 6.0









#25
2-Butanone
Concen: 99.308 ug/l
RT: 7.483 min Scan# 941
Delta R.T. -0.005 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 43 Resp: 117674
Ion Ratio Lower Upper
43 100
72 27.5 18.2 27.2#



#26
2,2-Dichloropropane
Concen: 20.385 ug/l
RT: 7.489 min Scan# 941
Delta R.T. -0.005 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion: 77 Resp: 60874
Ion Ratio Lower Upper
77 100
97 21.5 10.3 30.9









#27
cis-1,2-Dichloroethene
Concen: 19.311 ug/l
RT: 7.489 min Scan# 941
Delta R.T. 0.001 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 96 Resp: 39763
Ion Ratio Lower Upper
96 100
61 150.2 0.0 329.6
98 65.9 0.0 128.6



#28
Bromochloromethane
Concen: 19.575 ug/l
RT: 7.812 min Scan# 996
Delta R.T. 0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion: 49 Resp: 26971
Ion Ratio Lower Upper
49 100
129 1.1 0.0 3.8
130 80.3 50.5 75.7#









#29
Tetrahydrofuran
Concen: 100.310 ug/l
RT: 7.841 min Scan# 1001
Delta R.T. 0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 42 Resp: 74830
Ion Ratio Lower Upper
42 100
72 46.0 30.0 45.0#
71 43.3 28.3 42.5#



#30
Chloroform
Concen: 19.362 ug/l
RT: 7.965 min Scan# 1022
Delta R.T. 0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion: 83 Resp: 65971
Ion Ratio Lower Upper
83 100
85 64.4 50.9 76.3









#31
Cyclohexane
Concen: 19.543 ug/l
RT: 8.253 min Scan# 1071
Delta R.T. -0.006 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion: 56 Resp: 56951




Ion	Ratio	Lower	Upper
56	100		
69	29.5	24.5	36.7
84	93.4	66.0	99.0




#32
1,1,1-Trichloroethane
Concen: 19.878 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. -0.006 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion: 97 Resp: 62884

Ion	Ratio	Lower	Upper
97	100		
99	64.1	52.0	78.0
61	47.1	42.1	63.1









#33
1,2-Dichloroethane-d4
Concen: 20.118 ug/l
RT: 8.577 min Scan# 1126
Delta R.T. -0.005 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 65 Resp: 44843
Ion Ratio Lower Upper
65 100
67 50.8 0.0 104.4



#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 9.100 min Scan# 1215
Delta R.T. -0.006 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion: 114 Resp: 267615
Ion Ratio Lower Upper
114 100
63 19.8 0.0 44.6
88 15.6 0.0 31.4









#35
Dibromofluoromethane
Concen: 20.109 ug/l
RT: 8.165 min Scan# 1056
Delta R.T. 0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion:113 Resp: 34532
Ion Ratio Lower Upper
113 100
111 101.5 82.4 123.6
192 19.2 14.9 22.3



#36
1,1-Dichloropropene
Concen: 19.861 ug/l
RT: 8.371 min Scan# 1091
Delta R.T. -0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion: 75 Resp: 47023
Ion Ratio Lower Upper
75 100
110 34.2 16.3 48.9
77 29.7 24.6 37.0





#37

Ethyl Acetate

Concen: 18.579 ug/l

RT: 7.559 min Scan# 91

Delta R.T. -0.006 min

Lab File: VN082815.D

Acq: 08 Jul 2024 11:34

Instrument :

MSVOA_N

ClientSampleId :

VSTDIC020





Tgt Ion: 43 Resp: 44783

Ion Ratio Lower Upper

43 100

61 13.6 9.3 13.9

70 11.0 7.0 10.6#



#38

Carbon Tetrachloride

Concen: 19.544 ug/l

RT: 8.365 min Scan# 1090

Delta R.T. 0.000 min

Lab File: VN082815.D

Acq: 08 Jul 2024 11:34




Tgt Ion: 117 Resp: 55714


Ion Ratio Lower Upper

117 100

119 91.0 74.9 112.3

121 30.6 24.3 36.5









#39
Methylcyclohexane
Concen: 19.490 ug/l
RT: 9.600 min Scan# 1130
Delta R.T. -0.006 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 83 Resp: 49844
Ion Ratio Lower Upper
83 100
55 74.4 66.4 99.6
98 48.3 40.7 61.1



#40
Benzene
Concen: 19.446 ug/l
RT: 8.606 min Scan# 1131
Delta R.T. 0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion: 78 Resp: 141624
Ion Ratio Lower Upper
78 100
77 23.7 19.0 28.4









#41
Methacrylonitrile
Concen: 20.494 ug/l
RT: 7.783 min Scan# 991
Delta R.T. 0.001 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion: 41 Resp: 26613




Ion	Ratio	Lower	Upper
41	100		
39	56.7	45.5	68.3
67	70.6	50.5	75.7
52	29.6	24.0	36.0




#42
1,2-Dichloroethane
Concen: 19.879 ug/l
RT: 8.677 min Scan# 1143
Delta R.T. 0.006 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion: 62 Resp: 52120

Ion	Ratio	Lower	Upper
62	100		
98	8.3	0.0	15.8









#43
Isopropyl Acetate
Concen: 20.889 ug/l
RT: 8.688 min Scan# 1145
Delta R.T. -0.006 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 43 Resp: 102312
Ion Ratio Lower Upper
43 100
61 21.3 17.8 26.6
87 10.8 8.2 12.2



#44
Trichloroethene
Concen: 19.970 ug/l
RT: 9.353 min Scan# 1258
Delta R.T. 0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion:130 Resp: 37261
Ion Ratio Lower Upper
130 100
95 92.0 0.0 197.8









#45
1,2-Dichloropropane
Concen: 20.362 ug/l
RT: 9.624 min Scan# 1304
Delta R.T. 0.001 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 63 Resp: 34621
Ion Ratio Lower Upper
63 100
65 29.5 24.4 36.6



#46
Dibromomethane
Concen: 19.605 ug/l
RT: 9.712 min Scan# 1319
Delta R.T. -0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion: 93 Resp: 25968
Ion Ratio Lower Upper
93 100
95 83.6 65.8 98.6
174 93.5 71.7 107.5









#47
Bromodichloromethane
Concen: 19.644 ug/l
RT: 9.888 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion:	83	Resp:	53763
Ion	Ratio	Lower	Upper
83	100		
85	62.8	51.1	76.7
127	8.2	7.2	10.8



#48
Methyl methacrylate
Concen: 19.927 ug/l
RT: 9.683 min Scan# 1314
Delta R.T. 0.001 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion:	41	Resp:	37663
Ion	Ratio	Lower	Upper
41	100		
69	85.7	56.3	84.5#
39	67.0	50.3	75.5









#49
1,4-Dioxane
Concen: 405.531 ug/l
RT: 9.700 min Scan# 11
Delta R.T. 0.006 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 88 Resp: 16625
Ion Ratio Lower Upper
88 100
43 28.6 27.8 41.8
58 67.0 59.4 89.0



#50
Toluene-d8
Concen: 19.926 ug/l
RT: 10.571 min Scan# 1465
Delta R.T. 0.001 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion: 98 Resp: 126378
Ion Ratio Lower Upper
98 100
100 67.1 51.5 77.3









#51
4-Methyl-2-Pentanone
Concen: 100.562 ug/l
RT: 10.447 min Scan# 1444
Delta R.T. 0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 43 Resp: 240308
Ion Ratio Lower Upper
43 100
58 39.2 28.2 42.4



#52
Toluene
Concen: 20.134 ug/l
RT: 10.630 min Scan# 1475
Delta R.T. -0.005 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion: 92 Resp: 91596
Ion Ratio Lower Upper
92 100
91 171.6 139.4 209.0









#53
t-1,3-Dichloropropene
Concen: 20.199 ug/l
RT: 10.835 min Scan# 11
Delta R.T. 0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 75 Resp: 56950
Ion Ratio Lower Upper
75 100
77 30.8 25.7 38.5



#54
cis-1,3-Dichloropropene
Concen: 19.745 ug/l
RT: 10.312 min Scan# 1421
Delta R.T. -0.006 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion: 75 Resp: 58926
Ion Ratio Lower Upper
75 100
77 33.4 24.3 36.5
39 52.6 50.5 75.7









#55
1,1,2-Trichloroethane
Concen: 20.375 ug/l
RT: 11.018 min Scan# 1541
Delta R.T. -0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020

Tgt Ion: 97 Resp: 34086




Ion	Ratio	Lower	Upper
97	100		
83	84.7	71.4	107.0
85	54.7	45.8	68.6
99	63.7	52.0	78.0




#56
Ethyl methacrylate
Concen: 20.647 ug/l
RT: 10.877 min Scan# 1517
Delta R.T. 0.001 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion: 69 Resp: 53736

Ion	Ratio	Lower	Upper
69	100		
41	66.7	63.4	95.2
39	40.9	37.4	56.0









#57
1,3-Dichloropropane
Concen: 19.525 ug/l
RT: 11.165 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 76 Resp: 57819
Ion Ratio Lower Upper
76 100
78 33.5 25.6 38.4



#58
2-Chloroethyl Vinyl ether
Concen: 101.020 ug/l
RT: 10.159 min Scan# 1395
Delta R.T. -0.006 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion: 63 Resp: 134913
Ion Ratio Lower Upper
63 100
106 29.9 21.4 32.0









#59
2-Hexanone
Concen: 103.685 ug/l
RT: 11.200 min Scan# 1572
Delta R.T. 0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion: 43 Resp: 184214
Ion Ratio Lower Upper
43 100
58 55.0 24.6 73.8



#60
Dibromochloromethane
Concen: 20.029 ug/l
RT: 11.359 min Scan# 1599
Delta R.T. -0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion: 129 Resp: 42676
Ion Ratio Lower Upper
129 100
127 77.3 39.2 117.6









#61
1,2-Dibromoethane
Concen: 20.406 ug/l
RT: 11.471 min Scan# 1618
Delta R.T. 0.001 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion:107 Resp: 36253
Ion Ratio Lower Upper
107 100
109 92.3 73.7 110.5



#62
4-Bromofluorobenzene
Concen: 19.989 ug/l
RT: 12.847 min Scan# 1852
Delta R.T. 0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion: 95 Resp: 45541
Ion Ratio Lower Upper
95 100
174 77.6 0.0 159.2
176 76.6 0.0 147.6









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.871 min Scan# 1170
Delta R.T. 0.001 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion:117 Resp: 241012
Ion Ratio Lower Upper
117 100
82 54.5 47.5 71.3
119 32.2 25.6 38.4



#64
Tetrachloroethene
Concen: 19.516 ug/l
RT: 11.106 min Scan# 1556
Delta R.T. -0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion:164 Resp: 35092
Ion Ratio Lower Upper
164 100
166 127.8 101.4 152.0
129 99.2 75.7 113.5
131 92.1 73.8 110.6









#65
Chlorobenzene
Concen: 19.757 ug/l
RT: 11.894 min Scan# 111
Delta R.T. 0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion:112 Resp: 102037
Ion Ratio Lower Upper
112 100
114 31.3 26.2 39.2



#66
1,1,1,2-Tetrachloroethane
Concen: 20.107 ug/l
RT: 11.959 min Scan# 1701
Delta R.T. -0.005 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion:131 Resp: 36936
Ion Ratio Lower Upper
131 100
133 94.1 47.3 142.0
119 63.6 32.5 97.4









#67
Ethyl Benzene
Concen: 19.795 ug/l
RT: 11.965 min Scan# 1702
Delta R.T. 0.001 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 91 Resp: 171823
Ion Ratio Lower Upper
91 100
106 31.9 24.0 36.0



#68
m/p-Xylenes
Concen: 40.750 ug/l
RT: 12.077 min Scan# 1721
Delta R.T. 0.007 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion:106 Resp: 133538
Ion Ratio Lower Upper
106 100
91 203.0 166.1 249.1









#69
o-Xylene
Concen: 20.783 ug/l
RT: 12.400 min Scan# 1776
Delta R.T. 0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion:106 Resp: 65258
Ion Ratio Lower Upper
106 100
91 214.6 111.3 333.9



#70
Styrene
Concen: 20.451 ug/l
RT: 12.412 min Scan# 1778
Delta R.T. -0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion:104 Resp: 106102
Ion Ratio Lower Upper
104 100
78 51.4 41.6 62.4
103 55.1 44.0 66.0









#71
Bromoform
Concen: 20.604 ug/l
RT: 12.582 min Scan# 1807
Delta R.T. 0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion:	173	Resp:	29766
Ion Ratio	Lower	Upper	
173	100		
175	48.6	24.0	72.0
254	0.0	0.0	0.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.794 min Scan# 2013
Delta R.T. 0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion:	152	Resp:	110512
Ion Ratio	Lower	Upper	
152	100		
115	59.6	30.6	91.6
150	164.2	0.0	348.6









#73
Isopropylbenzene
Concen: 20.694 ug/l
RT: 12.694 min Scan# 1826
Delta R.T. -0.006 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion:105 Resp: 166182
Ion Ratio Lower Upper
105 100
120 26.6 12.9 38.6



#74
N-ethyl acetate
Concen: 20.803 ug/l
RT: 12.494 min Scan# 1792
Delta R.T. 0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion: 43 Resp: 69100
Ion Ratio Lower Upper
43 100
70 43.7 29.8 44.6
55 24.8 18.7 28.1
61 23.8 17.4 26.0









#75
1,1,2-Tetrachloroethane
Concen: 20.020 ug/l
RT: 12.941 min Scan# 1868
Delta R.T. 0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 83 Resp: 48149
Ion Ratio Lower Upper
83 100
131 10.8 5.0 14.9
85 64.7 31.9 95.7



#76
1,2,3-Trichloropropane
Concen: 31.090 ug/l
RT: 12.994 min Scan# 1877
Delta R.T. 0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion: 75 Resp: 67963
Ion Ratio Lower Upper
75 100
77 126.5 110.9 332.6









#77
Bromobenzene
Concen: 20.076 ug/l
RT: 12.982 min Scan# 1875
Delta R.T. 0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion:156 Resp: 41591
Ion Ratio Lower Upper
156 100
77 206.7 111.3 333.8
158 94.9 48.3 144.9



#78
n-propylbenzene
Concen: 20.398 ug/l
RT: 13.035 min Scan# 1884
Delta R.T. 0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion: 91 Resp: 188334
Ion Ratio Lower Upper
91 100
120 22.5 10.9 32.9









#79
2-Chlorotoluene
Concen: 19.957 ug/l
RT: 13.129 min Scan# 1900
Delta R.T. 0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion: 91 Resp: 117405
Ion Ratio Lower Upper
91 100
126 33.3 16.2 48.6



#80
1,3,5-Trimethylbenzene
Concen: 20.814 ug/l
RT: 13.176 min Scan# 1908
Delta R.T. 0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion:105 Resp: 135106
Ion Ratio Lower Upper
105 100
120 49.2 24.3 72.8









#81
trans-1,4-Dichloro-2-butene
Concen: 20.101 ug/l
RT: 12.735 min Scan# 1833
Delta R.T. -0.006 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Instrument : MSVOA_N
ClientSampleId : VSTDICC020

Tgt Ion: 75 Resp: 19473




Ion	Ratio	Lower	Upper
75	100		
53	106.9	88.4	132.6
89	46.3	36.3	54.5




#82
4-Chlorotoluene
Concen: 20.022 ug/l
RT: 13.224 min Scan# 1916
Delta R.T. 0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion: 91 Resp: 116588

Ion	Ratio	Lower	Upper
91	100		
126	33.9	16.2	48.6









#83
 tert-Butylbenzene
 Concen: 20.486 ug/l
 RT: 13.441 min Scan# 1953
 Delta R.T. 0.000 min
 Lab File: VN082815.D
 Acq: 08 Jul 2024 11:34

Instrument :
 MSVOA_N
 ClientSampleId :
 VSTDIC020

Tgt Ion:119 Resp: 117887




Ion	Ratio	Lower	Upper
119	100		
91	65.8	35.0	105.1
134	24.3	13.3	39.8




#84
 1,2,4-Trimethylbenzene
 Concen: 20.900 ug/l
 RT: 13.482 min Scan# 1960
 Delta R.T. 0.000 min
 Lab File: VN082815.D
 Acq: 08 Jul 2024 11:34

Tgt Ion:105 Resp: 138273

Ion	Ratio	Lower	Upper
105	100		
120	44.1	21.9	65.8









#85
sec-Butylbenzene
Concen: 20.607 ug/l
RT: 13.618 min Scan# 1983
Delta R.T. 0.001 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion:105 Resp: 160246
Ion Ratio Lower Upper
105 100
134 19.8 9.8 29.3



#86
p-Isopropyltoluene
Concen: 20.459 ug/l
RT: 13.729 min Scan# 2002
Delta R.T. 0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion:119 Resp: 132739
Ion Ratio Lower Upper
119 100
134 25.4 13.0 39.0
91 24.8 12.3 36.9





#87

1,3-Dichlorobenzene

Concen: 19.756 ug/l

RT: 13.735 min Scan# 2003

Delta R.T. 0.000 min

Lab File: VN082815.D

Acq: 08 Jul 2024 11:34

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020





Tgt Ion:146 Resp: 73937

Ion Ratio Lower Upper

146 100

111 41.0 20.6 61.9

148 63.3 31.8 95.3



#88

1,4-Dichlorobenzene

Concen: 19.901 ug/l

RT: 13.812 min Scan# 2016

Delta R.T. 0.001 min

Lab File: VN082815.D

Acq: 08 Jul 2024 11:34




Tgt Ion:146 Resp: 74739


Ion Ratio Lower Upper

146 100

111 41.9 20.1 60.3

148 65.0 32.2 96.6





#89

n-Butylbenzene

Concen: 20.756 ug/l

RT: 14.059 min Scan# 2058

Delta R.T. 0.001 min

Lab File: VN082815.D

Acq: 08 Jul 2024 11:34

Instrument :

MSVOA_N

ClientSampleId :

VSTDICC020





Tgt Ion: 91 Resp: 115371

Ion Ratio Lower Upper

91 100

92 52.0 26.1 78.3

134 24.5 12.1 36.3



#90

Hexachloroethane

Concen: 19.836 ug/l

RT: 14.335 min Scan# 2105

Delta R.T. 0.000 min

Lab File: VN082815.D




Acq: 08 Jul 2024 11:34


Tgt Ion: 117 Resp: 27232

Ion Ratio Lower Upper

117 100

201 70.4 35.8 107.3









#91
1,2-Dichlorobenzene
Concen: 19.936 ug/l
RT: 14.106 min Scan# 2066
Delta R.T. -0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDIC020


Tgt Ion:146 Resp: 72440
Ion Ratio Lower Upper
146 100
111 41.8 21.4 64.2
148 63.4 32.1 96.5



#92
1,2-Dibromo-3-Chloropropane
Concen: 19.849 ug/l
RT: 14.723 min Scan# 2171
Delta R.T. 0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion: 75 Resp: 11044
Ion Ratio Lower Upper
75 100
155 82.9 36.6 109.8
157 96.9 46.9 140.6









#93
1,2,4-Trichlorobenzene
Concen: 20.140 ug/l
RT: 15.394 min Scan# 2103
Delta R.T. -0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34




Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020


Tgt Ion:180 Resp: 39420
Ion Ratio Lower Upper
180 100
182 95.2 47.3 141.8
145 32.0 15.7 47.1



#94
Hexachlorobutadiene
Concen: 19.950 ug/l
RT: 15.500 min Scan# 2303
Delta R.T. -0.005 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion:225 Resp: 17310
Ion Ratio Lower Upper
225 100
223 64.3 31.9 95.7
227 62.9 32.5 97.5









#95
Naphthalene
Concen: 20.117 ug/l
RT: 15.641 min Scan# 2327
Delta R.T. -0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Instrument :
MSVOA_N
ClientSampleId :
VSTDICC020

Tgt Ion:	128	Resp:	134862
Ion	Ratio	Lower	Upper
128	100		
127	13.2	10.6	16.0
129	11.2	8.6	12.8



#96
1,2,3-Trichlorobenzene
Concen: 19.868 ug/l
RT: 15.841 min Scan# 2361
Delta R.T. -0.000 min
Lab File: VN082815.D
Acq: 08 Jul 2024 11:34

Tgt Ion:	180	Resp:	39984
Ion	Ratio	Lower	Upper
180	100		
182	95.2	48.9	146.8
145	33.8	16.8	50.4

