

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070824\
 Data File : VN082821.D
 Acq On : 08 Jul 2024 14:44
 Operator : JC\MD
 Sample : VN0708WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0708WBL01

Quant Time: Jul 08 17:42:51 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	95329	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	183112	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	177715	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	73786	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	74865	56.023	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	112.040%
35) Dibromofluoromethane	8.171	113	61841	52.631	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.260%
50) Toluene-d8	10.571	98	231708	53.392	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	106.780%
62) 4-Bromofluorobenzene	12.853	95	77933	49.993	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	99.980%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070824\
Data File : VN082821.D
Acq On : 08 Jul 2024 14:44
Operator : JC\MD
Sample : VN0708WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0708WBL01

Quant Time: Jul 08 17:42:51 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

