

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070824\
 Data File : VN082828.D
 Acq On : 08 Jul 2024 17:48
 Operator : JC\MD
 Sample : P3169-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 Storage-Blank-WATER-REF-5

Quant Time: Jul 09 06:06:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	96333	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	185421	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	191469	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	80369	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	77082	57.081	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 114.160%	
35) Dibromofluoromethane	8.171	113	63042	52.985	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 105.980%	
50) Toluene-d8	10.571	98	236501	53.818	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 107.640%	
62) 4-Bromofluorobenzene	12.853	95	84898	53.782	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 107.560%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070824\
Data File : VN082828.D
Acq On : 08 Jul 2024 17:48
Operator : JC\MD
Sample : P3169-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
Storage-Blank-WATER-REF-5

Quant Time: Jul 09 06:06:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

