

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070824\
 Data File : VN082829.D
 Acq On : 08 Jul 2024 18:13
 Operator : JC\MD
 Sample : P3169-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 Storage-Blank-WATER-REF-4

Quant Time: Jul 09 06:06:52 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	98904	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	187402	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	188133	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	78837	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	84656	61.060	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 122.120%	
35) Dibromofluoromethane	8.171	113	66951	55.676	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 111.360%	
50) Toluene-d8	10.571	98	241706	54.421	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 108.840%	
62) 4-Bromofluorobenzene	12.853	95	82341	51.611	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 103.220%	
Target Compounds						
16) Acetone	4.442	43	2488	5.097	ug/l	Qvalue # 81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070824\
Data File : VN082829.D
Acq On : 08 Jul 2024 18:13
Operator : JC\MD
Sample : P3169-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
Storage-Blank-WATER-REF-4

Quant Time: Jul 09 06:06:52 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

