

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070825\
 Data File : VN087297.D
 Acq On : 08 Jul 2025 14:18
 Operator : JC\MD
 Sample : VN0708WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0708WBL01

Quant Time: Jul 09 01:31:36 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N063025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 01 03:15:00 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	221705	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	437505	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	414800	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	206163	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	167125	60.062	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	120.120%
35) Dibromofluoromethane	8.171	113	148615	53.051	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.100%
50) Toluene-d8	10.565	98	541324	49.575	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.160%
62) 4-Bromofluorobenzene	12.847	95	182504	46.975	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	93.940%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070825\
Data File : VN087297.D
Acq On : 08 Jul 2025 14:18
Operator : JC\MD
Sample : VN0708WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0708WBL01

Quant Time: Jul 09 01:31:36 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N063025W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 01 03:15:00 2025
Response via : Initial Calibration

