

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070825\
 Data File : VN087300.D
 Acq On : 08 Jul 2025 15:36
 Operator : JC\MD
 Sample : Q2480-05 1000X
 Misc : 8.59g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GPX5

Quant Time: Jul 09 01:33:39 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N063025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 01 03:15:00 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	203934	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	381036	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	367407	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	171179	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	147457	57.612	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	115.220%
35) Dibromofluoromethane	8.171	113	125810	51.566	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.140%
50) Toluene-d8	10.565	98	467630	49.173	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	98.340%
62) 4-Bromofluorobenzene	12.847	95	159406	47.110	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	94.220%
Target Compounds						
					Qvalue	
84) 1,2,4-Trimethylbenzene	13.476	105	12439	1.347	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070825\
Data File : VN087300.D
Acq On : 08 Jul 2025 15:36
Operator : JC\MD
Sample : Q2480-05 1000X
Misc : 8.59g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GPX5

Quant Time: Jul 09 01:33:39 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N063025W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 01 03:15:00 2025
Response via : Initial Calibration

