

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070825\
 Data File : VN087303.D
 Acq On : 08 Jul 2025 16:39
 Operator : JC\MD
 Sample : Q2480-08 1000X
 Misc : 8.95g/5mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GPX8

Quant Time: Jul 09 01:34:40 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N063025W.M
 Quant Title : SW846 8260
 QLast Update : Tue Jul 01 03:15:00 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	327153	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	653023	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	368457	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	288480	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	272053	66.258	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	132.520%#
35) Dibromofluoromethane	8.171	113	220034	52.623	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.240%
50) Toluene-d8	10.565	98	814042	49.947	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.900%
62) 4-Bromofluorobenzene	12.847	95	234633	40.461	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	80.920%
Target Compounds						
84) 1,2,4-Trimethylbenzene	13.482	105	16000	1.028	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070825\
Data File : VN087303.D
Acq On : 08 Jul 2025 16:39
Operator : JC\MD
Sample : Q2480-08 1000X
Misc : 8.95g/5mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GPX8

Quant Time: Jul 09 01:34:40 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N063025W.M
Quant Title : SW846 8260
QLast Update : Tue Jul 01 03:15:00 2025
Response via : Initial Calibration

