


Data Path : Z:\voasrv\HPCHEM1\MSVOA N\Data\VN070919\
 Data File : VN056620.D
 Acq On : 9 Jul 2019 9:18
 Operator : JC/SP
 Sample : BFB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT3.P

Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\624N070319W.M
 Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 Last Update : Wed Jul 03 11:05:14 2019



AutoFind: Scans 3303, 3304, 3305; Background Corrected with Scan 3291

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	17.0	8128	PASS
75	95	30	60	45.6	21784	PASS
95	95	100	100	100.0	47776	PASS
96	95	5	9	7.0	3352	PASS
173	174	0.00	2	1.0	416	PASS
174	95	50	100	85.8	41002	PASS
175	174	5	9	7.0	2886	PASS
176	174	95	101	95.5	39162	PASS
177	176	5	9	6.9	2718	PASS