

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070920\
 Data File : VN062475.D
 Acq On : 9 Jul 2020 23:27
 Operator : JC/MD
 Sample : PB130035ZHE#06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB130035ZHE#06

Quant Time: Jul 10 02:11:07 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 10 00:55:52 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	204699	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	376950	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	362728	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	134459	50.00	ug/l	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
33) 1,2-Dichloroethane-d4	7.99	65	158000	50.18	ug/l	0.00
Spiked Amount			Recovery	=	100.36%	
35) Dibromofluoromethane	7.55	113	119201	51.16	ug/l	0.00
Spiked Amount			Recovery	=	102.32%	
50) Toluene-d8	10.06	98	486369	52.88	ug/l	0.00
Spiked Amount			Recovery	=	105.76%	
62) 4-Bromofluorobenzene	12.38	95	175235	52.13	ug/l	0.00
Spiked Amount			Recovery	=	104.26%	


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
11) Tert butyl alcohol	4.74	59	133	0.255	ug/l #	74
16) Acetone	3.78	43	14044	10.506	ug/l	98
18) Methyl Acetate	4.29	43	5734	1.719	ug/l #	87
20) Methylene Chloride	4.50	84	20460	7.159	ug/l	93
25) 2-Butanone	6.80	43	2144	1.098	ug/l #	84
29) Tetrahydrofuran	7.18	42	671	0.511	ug/l #	39
31) Cyclohexane	7.63	56	4992	1.044	ug/l #	1
36) 1,1-Dichloropropene	7.63	75	17845	4.699	ug/l #	52
38) Carbon Tetrachloride	7.63	117	20194	5.915	ug/l #	15
41) Methacrylonitrile	7.18	41	508	0.274	ug/l #	19
43) Isopropyl Acetate	8.13	43	70693	10.077	ug/l	93
45) 1,2-Dichloropropane	9.24	63	1481	0.456	ug/l #	44
48) Methyl methacrylate	9.14	41	4454	1.294	ug/l #	17
51) 4-Methyl-2-Pentanone	9.95	43	7024	1.676	ug/l	98
54) cis-1,3-Dichloropropene	9.87	75	27162	5.605	ug/l #	60
56) Ethyl methacrylate	10.52	69	1027	1.924	ug/l #	1
57) 1,3-Dichloropropane	10.74	76	5449	1.160	ug/l #	43
59) 2-Hexanone	10.74	43	78864	25.927	ug/l #	54
71) Bromoform	12.38	173	1158	0.608	ug/l #	1
74) N-amyl acetate	12.00	43	2891	0.571	ug/l #	48
76) 1,2,3-Trichloropropane	12.37	75	91646	27.779	ug/l #	100
80) 1,3,5-Trimethylbenzene	12.71	105	3228	0.378	ug/l	95
81) trans-1,4-Dichloro-2-buten	12.37	75	91646	74.352	ug/l #	10
95) Naphthalene	15.11	128	750	Below Cal	#	70


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070920\
Data File : VN062475.D
Acq On : 9 Jul 2020 23:27
Operator : JC/MD
Sample : PB130035ZHE#06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB130035ZHE#06

Quant Time: Jul 10 02:11:07 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 10 00:55:52 2020
Response via : Initial Calibration





#1

Pentafluorobenzene

Concen: 50.000 ug/l

RT: 7.63 min Scan# 1820

Delta R.T. 0.00 min

Lab File: VN062475.D

Acq: 9 Jul 2020 23:27

Instrument :

MSVOA_N

ClientSampleId :


PB130035ZHE#06

Tot Ion:168 Resp: 204699

Ion Ratio Lower Upper




168 100

99 61.6 48.8 73.2



Abundance Ion 167.90 (167.60 to 168.60): V

Ion 98.90 (98.60 to 99.60): VNO



#11

Tert butyl alcohol

Concen: 0.255 ug/l

RT: 4.74 min Scan# 922

Delta R.T. 0.00 min

Lab File: VN062475.D


Acq: 9 Jul 2020 23:27

Tgt Ion: 59 Resp: 133

Ion Ratio Lower Upper



59 100


57 0.0 7.8 11.6#



Abundance Ion 59.00 (58.70 to 59.70): VNO

Ion 57.00 (56.70 to 57.70): VNO





#16

Acetone

Concen: 10.506 ug/l

RT: 3.78 min Scan# 625

Delta R.T. 0.01 min

Lab File: VN062475.D

Acq: 9 Jul 2020 23:27

Instrument :

MSVOA_N

ClientSampleId :


PB130035ZHE#06

Tot Ion: 43 Resp: 14044

Ion Ratio Lower Upper




43 100

58 30.5 23.5 35.3



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 58.00 (57.70 to 58.70): VN0



#18

Methyl Acetate

Concen: 1.719 ug/l

RT: 4.29 min Scan# 782

Delta R.T. 0.01 min

Lab File: VN062475.D


Acq: 9 Jul 2020 23:27

Tgt Ion: 43 Resp: 5734

Ion Ratio Lower Upper



43 100


74 14.2 16.0 24.0#



Abundance Ion 43.00 (42.70 to 43.70): VN0

Ion 74.00 (73.70 to 74.70): VN0









#20
Methvlene Chloride
Concen: 7.159 ug/l
RT: 4.50 min Scan# 848
Delta R.T. 0.00 min
Lab File: VN062475.D
Acq: 9 Jul 2020 23:27

Instrument :
MSVOA_N
ClientSampleId :
PB130035ZHE#06

Tot Ion:	84	Resp:	20460
Ion	Ratio	Lower	Upper
84	100		
49	139.0	120.0	180.0
51	40.9	37.0	55.6
86	66.4	51.8	77.6




Abundance Ion 83.90 (83.60 to 84.60): VNC
Ion 48.90 (48.60 to 49.60): VNC
Ion 50.90 (50.60 to 51.60): VNC
Ion 85.90 (85.60 to 86.60): VNC






#25
2-Butanone
Concen: 1.098 ug/l
RT: 6.80 min Scan# 1562
Delta R.T. 0.01 min
Lab File: VN062475.D
Acq: 9 Jul 2020 23:27

Tgt Ion:	43	Resp:	2144
Ion	Ratio	Lower	Upper
43	100		
72	30.3	17.9	26.9#



Abundance Ion 43.00 (42.70 to 43.70): VNC
Ion 72.00 (71.70 to 72.70): VNC









#29
Tetrahydrofuran
Concen: 0.511 ug/l
RT: 7.18 min Scan# 1681
Delta R.T. 0.01 min
Lab File: VN062475.D
Acq: 9 Jul 2020 23:27




Instrument :
MSVOA_N
ClientSampleId :
PB130035ZHE#06


Tot Ion: 42 Resp: 671
Ion Ratio Lower Upper
42 100
72 0.0 29.1 43.7#
71 0.0 28.1 42.1#



#31
Cyclohexane
Concen: 1.044 ug/l
RT: 7.63 min Scan# 1820
Delta R.T. 0.01 min
Lab File: VN062475.D
Acq: 9 Jul 2020 23:27

Tgt Ion: 56 Resp: 4992
Ion Ratio Lower Upper
56 100
69 187.0 24.5 36.7#
84 133.2 57.4 86.2#





#33

1,2-Dichloroethane-d4

Concen: 50.182 ug/l

RT: 7.99 min Scan# 1933

Delta R.T. 0.00 min

Lab File: VN062475.D

Acq: 9 Jul 2020 23:27

Instrument :

MSVOA_N

ClientSampleId :


PB130035ZHE#06

Tot Ion: 65 Resp: 158000

Ion Ratio Lower Upper




65 100

67 54.0 0.0 107.6



Abundance Ion 64.90 (64.60 to 65.60): VN062475.D (-1840) (-)

Ion 66.90 (66.60 to 67.60): VN062475.D (-1840) (-)



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.55 min Scan# 2108

Delta R.T. 0.00 min

Lab File: VN062475.D

Acq: 9 Jul 2020 23:27


Tgt Ion: 114 Resp: 376950

Ion Ratio Lower Upper

114 100

63 24.1 0.0 48.4



88 17.0 0.0 33.8




Abundance Ion 113.90 (113.60 to 114.60): VN062475.D (-2015) (-)

Ion 63.00 (62.70 to 63.70): VN062475.D (-2015) (-)

Ion 87.90 (87.60 to 88.60): VN062475.D (-2015) (-)









#35
 Dibromofluoromethane
 Concen: 51.158 ug/l
 RT: 7.55 min Scan# 1797
 Delta R.T. 0.00 min
 Lab File: VN062475.D
 Acq: 9 Jul 2020 23:27




Instrument :
 MSVOA_N
 ClientSampleId :
 PB130035ZHE#06


Tot Ion:113 Resp: 119201
 Ion Ratio Lower Upper
 113 100
 111 103.3 83.1 124.7
 192 18.0 14.1 21.1



#36
 1,1-Dichloropropene
 Concen: 4.699 ug/l
 RT: 7.63 min Scan# 1820
 Delta R.T. -0.13 min
 Lab File: VN062475.D
 Acq: 9 Jul 2020 23:27

Tgt Ion: 75 Resp: 17845
 Ion Ratio Lower Upper
 75 100
 110 8.2 15.4 46.2#
 77 0.0 24.1 36.1#





#38

Carbon Tetrachloride

Concen: 5.915 ug/l

RT: 7.63 min Scan# 1820

Delta R.T. -0.11 min

Lab File: VN062475.D

Acq: 9 Jul 2020 23:27

Instrument :

MSVOA_N

ClientSampleId :

PB130035ZHE#06


Tot Ion:117 Resp: 20194

Ion Ratio Lower Upper

117 100

119 4.8 77.2 115.8#




121 0.0 24.8 37.2#



Abundance Ion 116.80 (116.50 to 117.50): V

Ion 118.80 (118.50 to 119.50): V

Ion 120.80 (120.50 to 121.50): V



#41

Methacrylonitrile

Concen: 0.274 ug/l

RT: 7.18 min Scan# 1682

Delta R.T. 0.06 min

Lab File: VN062475.D

Acq: 9 Jul 2020 23:27

Tgt Ion: 41 Resp: 508


Ion Ratio Lower Upper

41 100

39 0.0 50.2 75.2#

67 0.0 64.2 96.2#

52 0.0 28.0 42.0#






Abundance Ion 41.00 (40.70 to 41.70): V

Ion 39.00 (38.70 to 39.70): V

Ion 67.00 (66.70 to 67.70): V

Ion 52.00 (51.70 to 52.70): V









#43
Isopropyl Acetate
Concen: 10.077 ug/l
RT: 8.13 min Scan# 1976
Delta R.T. 0.00 min
Lab File: VN062475.D
Acq: 9 Jul 2020 23:27




Instrument :
MSVOA_N
ClientSampleId :
PB130035ZHE#06


Tot Ion: 43 Resp: 70693
Ion Ratio Lower Upper
43 100
61 19.6 19.3 28.9
87 10.5 8.6 13.0



#45
1,2-Dichloropropane
Concen: 0.456 ug/l
RT: 9.24 min Scan# 2322
Delta R.T. 0.15 min
Lab File: VN062475.D
Acq: 9 Jul 2020 23:27

Tgt Ion: 63 Resp: 1481
Ion Ratio Lower Upper
63 100
65 0.0 24.6 36.8#





#48

Methvl methacrylate

Concen: 1.294 ug/l

RT: 9.14 min Scan# 2292

Delta R.T. -0.02 min

Lab File: VN062475.D

Acq: 9 Jul 2020 23:27

Instrument :

MSVOA_N

ClientSampleId :

PB130035ZHE#06


Tot Ion: 41 Resp: 4454

Ion Ratio Lower Upper

41 100

69 0.0 60.8 91.2#




39 0.0 43.5 65.3#



Abundance Ion 41.00 (40.70 to 41.70): VN062401.D (-2287) (-)

Ion 69.00 (68.70 to 69.70): VN062401.D (-2287) (-)

Ion 39.00 (38.70 to 39.70): VN062401.D (-2287) (-)



#50

Toluene-d8

Concen: 52.876 ug/l

RT: 10.06 min Scan# 2577

Delta R.T. 0.00 min

Lab File: VN062475.D


Acq: 9 Jul 2020 23:27

Tgt Ion: 98 Resp: 486369

Ion Ratio Lower Upper



98 100


100 63.3 51.5 77.3



Abundance Ion 98.00 (97.70 to 98.70): VN062401.D (-2562) (-)

Ion 100.00 (99.70 to 100.70): VN062401.D (-2562) (-)









#51
 4-Methyl-2-Pentanone
 Concen: 1.676 ug/l
 RT: 9.95 min Scan# 2544
 Delta R.T. 0.00 min
 Lab File: VN062475.D
 Acq: 9 Jul 2020 23:27




Instrument :
 MSVOA_N
 ClientSampleId :
 PB130035ZHE#06


Tot Ion: 43 Resp: 7024
 Ion Ratio Lower Upper
 43 100
 58 35.3 29.0 43.6



#54
 cis-1,3-Dichloropropene
 Concen: 5.605 ug/l
 RT: 9.87 min Scan# 2519
 Delta R.T. 0.06 min
 Lab File: VN062475.D
 Acq: 9 Jul 2020 23:27

Tgt Ion: 75 Resp: 27162
 Ion Ratio Lower Upper
 75 100
 77 0.0 24.5 36.7#
 39 32.0 43.4 65.0#





#56

Ethyl methacrylate

Concen: 1.924 ug/l

RT: 10.52 min Scan# 2720

Delta R.T. 0.12 min

Lab File: VN062475.D

Acq: 9 Jul 2020 23:27

Instrument :

MSVOA_N

ClientSampleId :

PB130035ZHE#06


Tot Ion: 69 Resp: 1027

Ion Ratio Lower Upper

69 100

41 5713.9 59.7 89.5#




39 2244.2 30.6 45.8#



Abundance Ion 69.00 (68.70 to 69.70): VN062401.D (-2669) (-)

Ion 41.00 (40.70 to 41.70): VN062401.D (-2669) (-)

Ion 39.00 (38.70 to 39.70): VN062401.D (-2669) (-)



#57

1,3-Dichloropropane

Concen: 1.160 ug/l

RT: 10.74 min Scan# 2788

Delta R.T. 0.06 min

Lab File: VN062475.D


Acq: 9 Jul 2020 23:27

Tgt Ion: 76 Resp: 5449

Ion Ratio Lower Upper



76 100


78 0.0 25.6 38.4#



Abundance Ion 75.90 (75.60 to 76.60): VN062401.D (-2755) (-)

Ion 77.90 (77.60 to 78.60): VN062401.D (-2755) (-)









#59
2-Hexanone
Concen: 25.927 ug/l
RT: 10.74 min Scan# 2788
Delta R.T. 0.02 min
Lab File: VN062475.D
Acq: 9 Jul 2020 23:27




Instrument :
MSVOA_N
ClientSampleId :
PB130035ZHE#06


Tot Ion: 43 Resp: 78864
Ion Ratio Lower Upper
43 100
58 18.2 25.1 75.2#



#62
4-Bromofluorobenzene
Concen: 52.132 ug/l
RT: 12.38 min Scan# 3298
Delta R.T. 0.00 min
Lab File: VN062475.D
Acq: 9 Jul 2020 23:27

Tgt Ion: 95 Resp: 175235
Ion Ratio Lower Upper
95 100
174 71.8 0.0 141.6
176 68.3 0.0 137.2









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.38 min Scan# 2988
Delta R.T. 0.00 min
Lab File: VN062475.D
Acq: 9 Jul 2020 23:27

Instrument :
MSVOA_N
ClientSampleId :
PB130035ZHE#06

Tot Ion:117 Resp: 362728
Ion Ratio Lower Upper
117 100
82 61.8 49.6 74.4
119 31.5 25.4 38.0




Abundance Ion 116.90 (116.60 to 117.60): V
Ion 82.00 (81.70 to 82.70): V
Ion 118.90 (118.60 to 119.60): V






#71
Bromoform
Concen: 0.608 ug/l
RT: 12.38 min Scan# 3298
Delta R.T. 0.28 min
Lab File: VN062475.D
Acq: 9 Jul 2020 23:27

Tgt Ion:173 Resp: 1158
Ion Ratio Lower Upper
173 100
175 677.8 23.7 71.1#
254 0.0 0.0 0.0



Abundance Ion 172.70 (172.40 to 173.40): V
Ion 174.70 (174.40 to 175.40): V
Ion 254.40 (254.10 to 255.10): V





#72

1,4-Dichlorobenzene-d4

Concen: 50.000 ug/l

RT: 13.32 min Scan# 3591

Delta R.T. 0.00 min

Lab File: VN062475.D

Acq: 9 Jul 2020 23:27

Instrument :

MSVOA_N

ClientSampleId :

PB130035ZHE#06


Tot Ion:152 Resp: 134459

Ion Ratio Lower Upper

152 100

115 62.5 31.4 94.2




150 156.5 0.0 347.4



Abundance Ion 151.90 (151.60 to 152.60): V

Ion 114.90 (114.60 to 115.60): V

Ion 149.90 (149.60 to 150.60): V



#74

N-ethyl acetate

Concen: 0.571 ug/l

RT: 12.00 min Scan# 3182

Delta R.T. -0.04 min

Lab File: VN062475.D

Acq: 9 Jul 2020 23:27

Tgt Ion: 43 Resp: 2891


Ion Ratio Lower Upper

43 100

70 0.0 30.0 45.0#

55 3.4 18.2 27.4#

61 0.0 17.7 26.5#






Abundance Ion 43.00 (42.70 to 43.70): V

Ion 70.00 (69.70 to 70.70): V

Ion 55.00 (54.70 to 55.70): V

Ion 61.00 (60.70 to 61.70): V





#76

1,2,3-Trichloropropane

Concen: 27.779 ug/l

RT: 12.37 min Scan# 3297

Delta R.T. -0.15 min

Lab File: VN062475.D

Acq: 9 Jul 2020 23:27

Instrument :

MSVOA_N

ClientSampleId :





PB130035ZHE#06

Tot Ion: 75 Resp: 91646

Ion Ratio Lower Upper

75 100

77 1.2 0.0 0.0#



#80

1,3,5-Trimethylbenzene

Concen: 0.378 ug/l

RT: 12.71 min Scan# 3401

Delta R.T. 0.00 min

Lab File: VN062475.D




Acq: 9 Jul 2020 23:27


Tgt Ion: 105 Resp: 3228

Ion Ratio Lower Upper

105 100

120 43.7 23.4 70.2





#81

trans-1,4-Dichloro-2-butene

Concen: 74.352 ug/l

RT: 12.37 min Scan# 3297

Delta R.T. 0.10 min

Lab File: VN062475.D

Acq: 9 Jul 2020 23:27

Instrument :

MSVOA_N

ClientSampleId :

PB130035ZHE#06


Tot Ion: 75 Resp: 91646

Ion Ratio Lower Upper

75 100

53 0.0 79.2 118.8#




89 0.0 34.6 52.0#



Abundance Ion 74.90 (74.60 to 75.60): VN062475.D (-3173) (-)

Ion 53.00 (52.70 to 53.70): VN062475.D (-3173) (-)

Ion 88.90 (88.60 to 89.60): VN062475.D (-3173) (-)



#95

Naphthalene

Concen: Below Cal

RT: 15.11 min Scan# 4148

Delta R.T. 0.01 min

Lab File: VN062475.D

Acq: 9 Jul 2020 23:27


Tgt Ion: 128 Resp: 750

Ion Ratio Lower Upper

128 100

127 0.0 10.2 15.2#

129 0.0 8.6 12.8#



Abundance Ion 128.00 (127.70 to 128.70): VN062475.D (-4052) (-)

Ion 127.00 (126.70 to 127.70): VN062475.D (-4052) (-)

Ion 129.00 (128.70 to 129.70): VN062475.D (-4052) (-)

