

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070920\
 Data File : VN062453.D
 Acq On : 9 Jul 2020 13:29
 Operator : JC/MD
 Sample : L3215-19DL 20X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-6DL

Manual Integrations
 APPROVED

MMDadoda
 7/10/2020 9:23:56 AM

Quant Time: Jul 09 14:06:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 09 02:12:34 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	209185	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	377290	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	358894	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	147356	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	162611	50.54	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.08%	
35) Dibromofluoromethane	7.55	113	118147	50.66	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.32%	
50) Toluene-d8	10.06	98	485214	52.70	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.40%	
62) 4-Bromofluorobenzene	12.38	95	185278	55.07	ug/l	0.00
Spiked Amount 50.000			Recovery	=	110.14%	
Target Compounds						
					Qvalue	
39) Methylcyclohexane	9.05	83	25238	6.219	ug/l	99
40) Benzene	8.00	78	535378	47.789	ug/l	100
52) Toluene	10.12	92	132811	20.524	ug/l	98
67) Ethyl Benzene	11.48	91	323543	24.711	ug/l	100
68) m/p-Xylenes	11.59	106	704694	152.036	ug/l	98
69) o-Xylene	11.92	106	237139	52.547	ug/l	98
73) Isopropylbenzene	12.22	105	18243	1.600	ug/l	99
78) n-propylbenzene	12.57	91	28226	2.119	ug/l	98
80) 1,3,5-Trimethylbenzene	12.71	105	152336	16.297	ug/l	99
84) 1,2,4-Trimethylbenzene	13.01	105	574086	61.787	ug/l	100
86) p-Isopropyltoluene	13.26	119	2661m	0.293	ug/l	
95) Naphthalene	15.10	128	70273	13.441	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070920\
Data File : VN062453.D
Acq On : 9 Jul 2020 13:29
Operator : JC/MD
Sample : L3215-19DL 20X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-6DL

Manual Integrations
APPROVED

MMDadoda
7/10/2020 9:23:56 AM

Quant Time: Jul 09 14:06:42 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
Quant Title : SW846 8260
QLast Update : Thu Jul 09 02:12:34 2020
Response via : Initial Calibration

