

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070920\
 Data File : VN062464.D
 Acq On : 9 Jul 2020 19:01
 Operator : JC/MD
 Sample : L3192-08
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-MW180-FB-20200630

Quant Time: Jul 10 09:13:21 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 10 00:55:52 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	201390	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	363063	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	343198	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	124063	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	158337	51.12	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.24%	
35) Dibromofluoromethane	7.55	113	118371	52.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.48%	
50) Toluene-d8	10.06	98	464461	52.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.86%	
62) 4-Bromofluorobenzene	12.38	95	165031	50.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.94%	
Target Compounds						
16) Acetone	3.78	43	13603	10.344	ug/l	97
20) Methylene Chloride	4.51	84	1238	0.440	ug/l #	69
52) Toluene	10.12	92	39904	6.408	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070920\
Data File : VN062464.D
Acq On : 9 Jul 2020 19:01
Operator : JC/MD
Sample : L3192-08
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-MW180-FB-20200630

Quant Time: Jul 10 09:13:21 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 10 00:55:52 2020
Response via : Initial Calibration

