

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070920\
 Data File : VN062466.D
 Acq On : 9 Jul 2020 19:49
 Operator : JC/MD
 Sample : L3192-10RE
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-MW180-SOIL-604-608RE

Quant Time: Jul 10 02:00:08 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 10 00:55:52 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	208375	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	378069	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	366782	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	137055	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	164414	51.30	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.60%	
35) Dibromofluoromethane	7.55	113	118513	50.71	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.42%	
50) Toluene-d8	10.06	98	487974	52.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.78%	
62) 4-Bromofluorobenzene	12.38	95	178441	52.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.86%	
Target Compounds						
16) Acetone	3.78	43	18430	13.544	ug/l	93
18) Methyl Acetate	4.28	43	4889	1.440	ug/l #	71
20) Methylene Chloride	4.50	84	16092	5.531	ug/l	94
43) Isopropyl Acetate	8.13	43	80809	11.485	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070920\
Data File : VN062466.D
Acq On : 9 Jul 2020 19:49
Operator : JC/MD
Sample : L3192-10RE
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-MW180-SOIL-604-608RE

Quant Time: Jul 10 02:00:08 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 10 00:55:52 2020
Response via : Initial Calibration

