

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070920\
 Data File : VN062470.D
 Acq On : 9 Jul 2020 21:26
 Operator : JC/MD
 Sample : PB130035ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 27 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB130035ZHE#01

Quant Time: Jul 10 02:22:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 10 00:55:52 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	206827	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	376141	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	362283	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	136431	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	160096	50.32	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.64%	
35) Dibromofluoromethane	7.55	113	118659	51.03	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.06%	
50) Toluene-d8	10.06	98	487514	53.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.22%	
62) 4-Bromofluorobenzene	12.38	95	179797	53.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	107.20%	
Target Compounds						
16) Acetone	3.78	43	14827	10.978	ug/l	100
18) Methyl Acetate	4.28	43	6073	1.802	ug/l #	88
20) Methylene Chloride	4.50	84	18189	6.299	ug/l	98
43) Isopropyl Acetate	8.13	43	76717	10.959	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070920\
Data File : VN062470.D
Acq On : 9 Jul 2020 21:26
Operator : JC/MD
Sample : PB130035ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 27 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB130035ZHE#01

Quant Time: Jul 10 02:22:00 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 10 00:55:52 2020
Response via : Initial Calibration

