

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070920\
 Data File : VN062472.D
 Acq On : 9 Jul 2020 22:15
 Operator : JC/MD
 Sample : PB130035ZHE#03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB130035ZHE#03

Quant Time: Jul 10 02:25:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 10 00:55:52 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	208473	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	379470	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	367414	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	138260	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	160099	49.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.86%	
35) Dibromofluoromethane	7.55	113	119925	51.13	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.26%	
50) Toluene-d8	10.06	98	491128	53.04	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.08%	
62) 4-Bromofluorobenzene	12.38	95	180851	53.45	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.90%	
Target Compounds						
16) Acetone	3.78	43	13907	10.215	ug/l	94
18) Methyl Acetate	4.28	43	6123	1.802	ug/l #	70
20) Methylene Chloride	4.50	84	18788	6.455	ug/l	92
43) Isopropyl Acetate	8.13	43	74746	10.584	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070920\
Data File : VN062472.D
Acq On : 9 Jul 2020 22:15
Operator : JC/MD
Sample : PB130035ZHE#03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 29 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB130035ZHE#03

Quant Time: Jul 10 02:25:39 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 10 00:55:52 2020
Response via : Initial Calibration

