

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN070920\
 Data File : VN062474.D
 Acq On : 9 Jul 2020 23:03
 Operator : JC/MD
 Sample : PB130035ZHE#05
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB130035ZHE#05

Quant Time: Jul 10 02:29:39 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 10 00:55:52 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	201253	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	371816	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	362355	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	136936	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	159755	51.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.22%	
35) Dibromofluoromethane	7.55	113	117347	51.06	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.12%	
50) Toluene-d8	10.06	98	480971	53.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.02%	
62) 4-Bromofluorobenzene	12.38	95	173496	52.33	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.66%	
Target Compounds						
16) Acetone	3.79	43	13610	10.356	ug/l	94
18) Methyl Acetate	4.29	43	5927	1.807	ug/l #	82
20) Methylene Chloride	4.50	84	19898	7.081	ug/l	86
43) Isopropyl Acetate	8.13	43	75445	10.903	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN070920\
Data File : VN062474.D
Acq On : 9 Jul 2020 23:03
Operator : JC/MD
Sample : PB130035ZHE#05
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 31 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB130035ZHE#05

Quant Time: Jul 10 02:29:39 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 10 00:55:52 2020
Response via : Initial Calibration

