

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN070921\  
 Data File : VN067691.D  
 Acq On : 9 Jul 2021 17:28  
 Operator : JC/MD  
 Sample : IBLK  
 Misc : 5.00mL/MSVOA\_N/WATER  
 ALS Vial : 16 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 IBLK

Quant Time: Jul 10 05:25:19 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N070721W.M  
 Quant Title : SW846 8260  
 QLast Update : Wed Jul 07 16:18:53 2021  
 Response via : Initial Calibration

| Compound                   | R.T.   | QIon | Response | Conc   | Units | Dev(Min) |
|----------------------------|--------|------|----------|--------|-------|----------|
| -----                      |        |      |          |        |       |          |
| Internal Standards         |        |      |          |        |       |          |
| 1) Pentafluorobenzene      | 8.088  | 168  | 313240   | 50.000 | ug/l  | 0.00     |
| 34) 1,4-Difluorobenzene    | 8.971  | 114  | 583744   | 50.000 | ug/l  | 0.00     |
| 63) Chlorobenzene-d5       | 11.749 | 117  | 541558   | 50.000 | ug/l  | 0.00     |
| 72) 1,4-Dichlorobenzene-d4 | 13.678 | 152  | 196795   | 50.000 | ug/l  | 0.00     |

|                             |        |       |          |          |      |          |
|-----------------------------|--------|-------|----------|----------|------|----------|
| System Monitoring Compounds |        |       |          |          |      |          |
| 33) 1,2-Dichloroethane-d4   | 8.440  | 65    | 245168   | 51.789   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 61 - 141 | Recovery | =    | 103.580% |
| 35) Dibromofluoromethane    | 8.024  | 113   | 174216   | 49.397   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 69 - 133 | Recovery | =    | 98.800%  |
| 50) Toluene-d8              | 10.446 | 98    | 746066   | 50.538   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 65 - 126 | Recovery | =    | 101.080% |
| 62) 4-Bromofluorobenzene    | 12.734 | 95    | 237658   | 45.573   | ug/l | 0.00     |
| Spiked Amount               | 50.000 | Range | 58 - 135 | Recovery | =    | 91.140%  |

| Target Compounds | Qvalue |
|------------------|--------|
| -----            |        |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN070921\  
Data File : VN067691.D  
Acq On : 9 Jul 2021 17:28  
Operator : JC/MD  
Sample : IBLK  
Misc : 5.00mL/MSVOA\_N/WATER  
ALS Vial : 16 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
IBLK

Quant Time: Jul 10 05:25:19 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N070721W.M  
Quant Title : SW846 8260  
QLast Update : Wed Jul 07 16:18:53 2021  
Response via : Initial Calibration

