

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070921\
 Data File : VN067692.D
 Acq On : 9 Jul 2021 17:53
 Operator : JC/MD
 Sample : PB137589TB
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB137589TB

Manual Integrations
 APPROVED

MMDadoda
 7/12/2021 5:44:52 PM

Quant Time: Jul 10 05:25:39 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	315901	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	599370	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	646321	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.678	152	219596	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	246927	51.721	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 103.440%			
35) Dibromofluoromethane	8.024	113	176064	48.620	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 97.240%			
50) Toluene-d8	10.446	98	803646	53.019	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 106.040%			
62) 4-Bromofluorobenzene	12.734	95	271995	50.797	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 101.600%			
Target Compounds					Qvalue	
16) Acetone	4.296	43	33673	22.415	ug/l	96
20) Methylene Chloride	5.085	84	14725m	3.552	ug/l	
43) Isopropyl Acetate	8.566	43	47015	4.955	ug/l #	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070921\
Data File : VN067692.D
Acq On : 9 Jul 2021 17:53
Operator : JC/MD
Sample : PB137589TB
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB137589TB

Manual Integrations
APPROVED

MMDadoda
7/12/2021 5:44:52 PM

Quant Time: Jul 10 05:25:39 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

