

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070921\
 Data File : VN067694.D
 Acq On : 9 Jul 2021 18:42
 Operator : JC/MD
 Sample : M2992-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 33287

Manual Integrations
 APPROVED

MMDadoda
 7/12/2021 5:44:55 PM

Quant Time: Jul 10 05:26:17 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jul 07 16:18:53 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.088	168	394591	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.971	114	719777	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.747	117	654814	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.675	152	345620	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.440	65	306880	51.460	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 102.920%			
35) Dibromofluoromethane	8.024	113	223356	51.361	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 102.720%			
50) Toluene-d8	10.443	98	902627	49.588	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 99.180%			
62) 4-Bromofluorobenzene	12.734	95	311769	48.485	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 96.980%			
Target Compounds						Qvalue
11) Tert butyl alcohol	5.390	59	190790	222.638	ug/l	99
16) Acetone	4.288	43	1053773	561.574	ug/l	96
18) Methyl Acetate	4.865	43	6870	1.303	ug/l	98
25) 2-Butanone	7.324	43	37529661m	12044.428	ug/l	
29) Tetrahydrofuran	7.691	42	125075	61.039	ug/l	90
40) Benzene	8.461	78	37008	1.849	ug/l	97
70) Styrene	12.296	104	642583	44.122	ug/l #	85
95) Naphthalene	15.507	128	19087	4.072	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070921\
Data File : VN067694.D
Acq On : 9 Jul 2021 18:42
Operator : JC/MD
Sample : M2992-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
33287

Manual Integrations
APPROVED

MMDadoda
7/12/2021 5:44:55 PM

Quant Time: Jul 10 05:26:17 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070721W.M
Quant Title : SW846 8260
QLast Update : Wed Jul 07 16:18:53 2021
Response via : Initial Calibration

