

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070924\
 Data File : VN082837.D
 Acq On : 09 Jul 2024 14:09
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jul 10 01:20:26 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	101613	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	190318	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	178201	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	72541	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	76633	53.800	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.600%
35) Dibromofluoromethane	8.171	113	62354	51.059	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	102.120%
50) Toluene-d8	10.571	98	215712	47.824	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.640%
62) 4-Bromofluorobenzene	12.853	95	70786	43.689	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	87.380%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070924\
Data File : VN082837.D
Acq On : 09 Jul 2024 14:09
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jul 10 01:20:26 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

