

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070924\
 Data File : VN082842.D
 Acq On : 09 Jul 2024 16:21
 Operator : JC\MD
 Sample : P3173-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-189-GW-763-765

Quant Time: Jul 10 01:23:10 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	97205	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	183488	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	179597	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	73666	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	74334	54.552	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.100%	
35) Dibromofluoromethane	8.171	113	61039	51.842	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 103.680%	
50) Toluene-d8	10.571	98	207992	47.829	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 95.660%	
62) 4-Bromofluorobenzene	12.853	95	69348	44.394	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 88.780%	
Target Compounds						
16) Acetone	4.436	43	4042	8.425	ug/l	Qvalue 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070924\
Data File : VN082842.D
Acq On : 09 Jul 2024 16:21
Operator : JC\MD
Sample : P3173-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-189-GW-763-765

Quant Time: Jul 10 01:23:10 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

