

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070924\
 Data File : VN082844.D
 Acq On : 09 Jul 2024 17:10
 Operator : JC\MD
 Sample : P3173-05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BP-VPB-189-GW-778-780

Quant Time: Jul 10 01:23:57 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	94427	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	178497	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	175984	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	72017	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	76040	57.446	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.900%
35) Dibromofluoromethane	8.177	113	61222	53.452	ug/l	0.01
Spiked Amount	50.000	Range	75 - 124	Recovery	=	106.900%
50) Toluene-d8	10.571	98	210261	49.702	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.400%
62) 4-Bromofluorobenzene	12.853	95	69626	45.819	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	91.640%
Target Compounds						
					Qvalue	
16) Acetone	4.442	43	6710	14.398	ug/l #	88
25) 2-Butanone	7.500	43	1644	2.336	ug/l #	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070924\
Data File : VN082844.D
Acq On : 09 Jul 2024 17:10
Operator : JC\MD
Sample : P3173-05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BP-VPB-189-GW-778-780

Quant Time: Jul 10 01:23:57 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

