

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070924\
 Data File : VN082845.D
 Acq On : 09 Jul 2024 17:35
 Operator : JC\MD
 Sample : P3170-05 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WATER-DRUMS

Quant Time: Jul 10 01:24:19 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	97320	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	182897	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	177745	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	70932	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	74412	54.545	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	109.100%
35) Dibromofluoromethane	8.171	113	62129	52.939	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	105.880%
50) Toluene-d8	10.570	98	212149	48.942	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	97.880%
62) 4-Bromofluorobenzene	12.853	95	67698	43.478	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	86.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN070924\
Data File : VN082845.D
Acq On : 09 Jul 2024 17:35
Operator : JC\MD
Sample : P3170-05 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WATER-DRUMS

Quant Time: Jul 10 01:24:19 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

