

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071019\
 Data File : VN056633.D
 Acq On : 10 Jul 2019 9:18
 Operator : JC/SP
 Sample : VSSTDICC0.1
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSSTDICC0.1

Quant Time: Jul 12 08:22:55 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071019SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 11 04:48:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	59044	5.00	ug/l	0.00
4) 1,4-Difluorobenzene	8.58	114	86564	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	65048	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.35	152	16143	5.00	ug/l	0.00

System Monitoring Compounds

3) 1,2-Dichloroethane-d4	0.00	65	0d	0.00	ug/l	
Spiked Amount	5.000	Range	81 - 118	Recovery	=	0.00%#
5) Dibromofluoromethane	0.00	113	0d	0.00	ug/l	
Spiked Amount	5.000	Range	80 - 119	Recovery	=	0.00%#
6) Toluene-d8	0.00	98	0d	0.00	ug/l	
Spiked Amount	5.000	Range	89 - 112	Recovery	=	0.00%#
7) 4-Bromofluorobenzene	0.00	95	0d	0.00	ug/l	
Spiked Amount	5.000	Range	85 - 114	Recovery	=	0.00%#

Target Compounds					Qvalue
2) Vinyl Chloride	2.17	62	584	0.100 ug/l #	62

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071019\
Data File : VN056633.D
Acq On : 10 Jul 2019 9:18
Operator : JC/SP
Sample : VSSTDICC0.1
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSSTDICC0.1

Quant Time: Jul 12 08:22:55 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071019SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Thu Jul 11 04:48:42 2019
Response via : Initial Calibration

