

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071019\
 Data File : VN056636.D
 Acq On : 10 Jul 2019 10:28
 Operator : JC/SP
 Sample : VSSTDICC002
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSSTDICC002

Quant Time: Jul 11 04:54:08 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071019SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 11 04:48:42 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	50787	5.00	ug/l	0.00
4) 1,4-Difluorobenzene	8.58	114	88254	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	57507	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.34	152	13993	5.00	ug/l	0.00

System Monitoring Compounds

3) 1,2-Dichloroethane-d4	8.02	65	12328	1.99	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	39.80%#
5) Dibromofluoromethane	7.59	113	9451	1.94	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	38.80%#
6) Toluene-d8	10.09	98	22865	2.05	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	41.00%#
7) 4-Bromofluorobenzene	12.39	95	7938	2.58	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	51.60%#

Target Compounds

	R.T.	QIon	Response	Conc	Units	Qvalue
2) Vinyl Chloride	2.17	62	9439	1.875	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071019\
Data File : VN056636.D
Acq On : 10 Jul 2019 10:28
Operator : JC/SP
Sample : VSSTDICC002
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSSTDICC002

Quant Time: Jul 11 04:54:08 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071019SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Thu Jul 11 04:48:42 2019
Response via : Initial Calibration

