

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071019\
 Data File : VN056641.D
 Acq On : 10 Jul 2019 13:15
 Operator : JC/SP
 Sample : 0.05PPB LOD MDL
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 0.05PPB LOD MDL

Quant Time: Jul 11 07:13:48 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071019SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 11 04:54:57 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	47980	5.00	ug/l	0.00
4) 1,4-Difluorobenzene	8.58	114	83543	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	55341	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.34	152	13372	5.00	ug/l	0.00

System Monitoring Compounds

3) 1,2-Dichloroethane-d4	8.01	65	28537	4.96	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	99.20%
5) Dibromofluoromethane	7.59	113	21860	4.96	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	99.20%
6) Toluene-d8	10.09	98	51867	4.86	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	97.20%
7) 4-Bromofluorobenzene	12.39	95	17809	4.66	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	93.20%

Target Compounds					Qvalue
2) Vinyl Chloride	2.17	62	213	0.047 ug/l #	43

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071019\
Data File : VN056641.D
Acq On : 10 Jul 2019 13:15
Operator : JC/SP
Sample : 0.05PPB LOD MDL
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
0.05PPB LOD MDL

Quant Time: Jul 11 07:13:48 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071019SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Thu Jul 11 04:54:57 2019
Response via : Initial Calibration

