

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071019\
 Data File : VN056644.D
 Acq On : 10 Jul 2019 14:47
 Operator : JC/SP
 Sample : K3636-02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SP19-JMW0961

Quant Time: Jul 11 07:18:03 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071019SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 11 04:54:57 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	50472	5.00	ug/l	0.00
4) 1,4-Difluorobenzene	8.58	114	86719	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	58924	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.35	152	14861	5.00	ug/l	0.00

System Monitoring Compounds

3) 1,2-Dichloroethane-d4	8.01	65	29642	4.90	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	98.00%
5) Dibromofluoromethane	7.59	113	22817	4.98	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	99.60%
6) Toluene-d8	10.09	98	53848	4.86	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	97.20%
7) 4-Bromofluorobenzene	12.40	95	19370	4.88	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	97.60%

Target Compounds					Qvalue
2) Vinyl Chloride	2.17	62	466	0.098 ug/l #	64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071019\
Data File : VN056644.D
Acq On : 10 Jul 2019 14:47
Operator : JC/SP
Sample : K3636-02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SP19-JMW0961

Quant Time: Jul 11 07:18:03 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071019SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Thu Jul 11 04:54:57 2019
Response via : Initial Calibration

