

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071019\
 Data File : VN056645.D
 Acq On : 10 Jul 2019 15:10
 Operator : JC/SP
 Sample : VN0710WBS02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0710WBS02

Quant Time: Jul 11 07:18:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071019SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 11 04:54:57 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	50099	5.00	ug/l	0.00
4) 1,4-Difluorobenzene	8.58	114	86245	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	56896	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.35	152	13749	5.00	ug/l	0.00

System Monitoring Compounds						
3) 1,2-Dichloroethane-d4	8.01	65	28160	4.69	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	93.80%
5) Dibromofluoromethane	7.59	113	21666	4.76	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	95.20%
6) Toluene-d8	10.09	98	51318	4.66	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	93.20%
7) 4-Bromofluorobenzene	12.40	95	17686	4.48	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	89.60%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071019\
Data File : VN056645.D
Acq On : 10 Jul 2019 15:10
Operator : JC/SP
Sample : VN0710WBS02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0710WBS02

Quant Time: Jul 11 07:18:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071019SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Thu Jul 11 04:54:57 2019
Response via : Initial Calibration

