

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071019\
 Data File : VN056649.D
 Acq On : 10 Jul 2019 16:46
 Operator : JC/SP
 Sample : K3636-09
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SP19-JMW0941

Quant Time: Jul 11 07:20:35 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071019SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 11 04:54:57 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.66	168	48090	5.00	ug/l	0.00
4) 1,4-Difluorobenzene	8.59	114	83826	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	55921	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.34	152	13457	5.00	ug/l	0.00

System Monitoring Compounds						
3) 1,2-Dichloroethane-d4	8.01	65	29560	5.12	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	102.40%
5) Dibromofluoromethane	7.58	113	22615	5.11	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	102.20%
6) Toluene-d8	10.09	98	53554	5.00	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	100.00%
7) 4-Bromofluorobenzene	12.39	95	18373	4.79	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	95.80%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071019\
Data File : VN056649.D
Acq On : 10 Jul 2019 16:46
Operator : JC/SP
Sample : K3636-09
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SP19-JMW0941

Quant Time: Jul 11 07:20:35 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071019SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Thu Jul 11 04:54:57 2019
Response via : Initial Calibration

