

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071019\
 Data File : VN056654.D
 Acq On : 10 Jul 2019 18:42
 Operator : JC/SP
 Sample : K3636-05MS
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SP19-JMW0991MS

Quant Time: Jul 11 07:23:32 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071019SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 11 04:54:57 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	48547	5.00	ug/l	0.00
4) 1,4-Difluorobenzene	8.58	114	85252	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	57281	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.35	152	14294	5.00	ug/l	0.00

System Monitoring Compounds

3) 1,2-Dichloroethane-d4	8.01	65	28915	4.97	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	99.40%
5) Dibromofluoromethane	7.59	113	22212	4.93	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	98.60%
6) Toluene-d8	10.09	98	52208	4.80	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	96.00%
7) 4-Bromofluorobenzene	12.40	95	18316	4.70	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	94.00%

Target Compounds

	R.T.	QIon	Response	Conc	Units	Qvalue
2) Vinyl Chloride	2.17	62	22513	4.929	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071019\
Data File : VN056654.D
Acq On : 10 Jul 2019 18:42
Operator : JC/SP
Sample : K3636-05MS
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SP19-JMW0991MS

Quant Time: Jul 11 07:23:32 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071019SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Thu Jul 11 04:54:57 2019
Response via : Initial Calibration

