

Method Path : Z:\V0ASRV\HPCHEM1\MSVOA N\METHODS\
 Method File : 82N071019SIM-VINYL.M
 Title : SW846 8260
 Last Update : Thu Jul 11 04:54:57 2019
 Response Via : Initial Calibration

Calibration Files

0.1 =VN056633.D 0.5 =VN056634.D 1 =VN056635.D
 5 =VN056637.D 8 =VN056638.D 2 =VN056636.D

	Compound	0.1	0.5	1	5	8	2	Avg	%RSD
1) I	Pentafluorobenzene	-----ISTD-----							
2) C	Vinyl Chloride	0.495	0.471	0.435	0.472	0.484	0.465	0.470	4.30#
3) S	1,2-Dichloroethan		0.623	0.585	0.586	0.598	0.607	0.600	2.64
4) I	1,4-Difluorobenzene	-----ISTD-----							
5) S	Dibromofluorometh		0.273	0.260	0.257	0.262	0.268	0.264	2.43
6) S	Toluene-d8		0.646	0.626	0.631	0.641	0.648	0.638	1.51
7) S	4-Bromofluorobenz		0.255	0.224	0.217	0.222	0.225	0.229	6.67
8) I	Chlorobenzene-d5	-----ISTD-----							
9) I	1,4-Dichlorobenzene-d	-----ISTD-----							

(#) = Out of Range