

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071019\
Data File : VN056639.D
Acq On : 10 Jul 2019 11:37
Operator : JC/SP
Sample : VSSTDICV005
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
ICVVN071019

Quant Time: Jul 11 05:00:39 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071019SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Thu Jul 11 04:54:57 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	5.000	5.000	0.0	98	0.00
2 C	Vinyl Chloride	5.000	5.038	-0.8#	98	0.00
3 S	1,2-Dichloroethane-d4	5.000	4.874	2.5	98	0.00
4 I	1,4-Difluorobenzene	5.000	5.000	0.0	97	0.00
5 S	Dibromofluoromethane	5.000	4.954	0.9	99	0.00
6 S	Toluene-d8	5.000	5.012	-0.2	99	0.00
7 S	4-Bromofluorobenzene	5.000	4.848	3.0	100	0.00
8 I	Chlorobenzene-d5	5.000	5.000	0.0	99	0.00
9 I	1,4-Dichlorobenzene-d4	5.000	5.000	0.0	101	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1