

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071020\
 Data File : VN062496.D
 Acq On : 10 Jul 2020 16:43
 Operator : JC/MD
 Sample : L3250-10 50X
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 DRUM-VAULT-WATER-COMP-2

Quant Time: Jul 11 02:38:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 10 00:55:52 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	188660	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	332741	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	317586	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	116419	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	144432	49.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.54%	
35) Dibromofluoromethane	7.55	113	109900	53.43	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.86%	
50) Toluene-d8	10.06	98	429527	52.90	ug/l	0.00
Spiked Amount 50.000			Recovery	=	105.80%	
62) 4-Bromofluorobenzene	12.38	95	151353	51.01	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.02%	
Target Compounds						
16) Acetone	3.79	43	7499	6.087	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071020\
Data File : VN062496.D
Acq On : 10 Jul 2020 16:43
Operator : JC/MD
Sample : L3250-10 50X
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
DRUM-VAULT-WATER-COMP-2

Quant Time: Jul 11 02:38:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 10 00:55:52 2020
Response via : Initial Calibration

