

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071020\
 Data File : VN062499.D
 Acq On : 10 Jul 2020 17:56
 Operator : JC/MD
 Sample : L3273-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-WATER-REF-4

Quant Time: Jul 11 02:50:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 10 00:55:52 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	189559	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	339756	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	326441	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	117868	50.00	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	147347	50.54	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.08%	
35) Dibromofluoromethane	7.55	113	108556	51.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	103.38%	
50) Toluene-d8	10.06	98	436837	52.69	ug/l	0.00
Spiked Amount	50.000		Recovery	=	105.38%	
62) 4-Bromofluorobenzene	12.38	95	155111	51.20	ug/l	0.00
Spiked Amount	50.000		Recovery	=	102.40%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071020\
Data File : VN062499.D
Acq On : 10 Jul 2020 17:56
Operator : JC/MD
Sample : L3273-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-WATER-REF-4

Quant Time: Jul 11 02:50:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 10 00:55:52 2020
Response via : Initial Calibration

