

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071020\
 Data File : VN062501.D
 Acq On : 10 Jul 2020 18:45
 Operator : JC/MD
 Sample : L3215-20
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FIELD-BLANK

Manual Integrations
 APPROVED

MMDadoda
 7/13/2020 11:19:36 AM

Quant Time: Jul 11 03:04:31 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
 Quant Title : SW846 8260
 QLast Update : Fri Jul 10 00:55:52 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.63	168	187733	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.55	114	334342	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.38	117	323056	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.32	152	118295	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.99	65	143482	49.69	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.38%	
35) Dibromofluoromethane	7.55	113	108275	52.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.78%	
50) Toluene-d8	10.06	98	433796	53.17	ug/l	0.00
Spiked Amount 50.000			Recovery	=	106.34%	
62) 4-Bromofluorobenzene	12.38	95	152384	51.11	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.22%	
Target Compounds						
16) Acetone	3.79	43	11517	9.394	ug/l	98
20) Methylene Chloride	4.51	84	2355m	0.898	ug/l	
52) Toluene	10.12	92	137251	23.935	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071020\
Data File : VN062501.D
Acq On : 10 Jul 2020 18:45
Operator : JC/MD
Sample : L3215-20
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FIELD-BLANK

Manual Integrations
APPROVED

MMDadoda
7/13/2020 11:19:36 AM

Quant Time: Jul 11 03:04:31 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N070820W.M
Quant Title : SW846 8260
QLast Update : Fri Jul 10 00:55:52 2020
Response via : Initial Calibration

