

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071023\
 Data File : VN078455.D
 Acq On : 10 Jul 2023 15:10
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Jul 10 16:20:14 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 10 16:13:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.235	168	348956	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.112	114	614741	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	514953	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	114189	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.588	65	234291	50.082	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	100.160%
35) Dibromofluoromethane	8.177	113	208260	49.906	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.820%
50) Toluene-d8	10.576	98	741089	47.010	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.020%
62) 4-Bromofluorobenzene	12.853	95	178242	36.707	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	73.420%#

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071023\
Data File : VN078455.D
Acq On : 10 Jul 2023 15:10
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Jul 10 16:20:14 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N071023W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 10 16:13:53 2023
Response via : Initial Calibration

