

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071024\
 Data File : VN082854.D
 Acq On : 10 Jul 2024 14:48
 Operator : JC\MD
 Sample : P3183-01 20X
 Misc : 5.06g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 N28660

Quant Time: Jul 11 03:48:45 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	88549	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	168834	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.871	117	174476	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	72486	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	77105	62.117	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	124.240%
35) Dibromofluoromethane	8.171	113	63809	58.899	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	117.800%
50) Toluene-d8	10.571	98	220336	55.065	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	110.120%
62) 4-Bromofluorobenzene	12.853	95	75107	52.254	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	104.500%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071024\
Data File : VN082854.D
Acq On : 10 Jul 2024 14:48
Operator : JC\MD
Sample : P3183-01 20X
Misc : 5.06g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
N28660

Quant Time: Jul 11 03:48:45 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

