

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071119\
 Data File : VN056659.D
 Acq On : 11 Jul 2019 9:26
 Operator : JC/SP
 Sample : VSSTDCCC005
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VSSTDCCC005

Quant Time: Jul 12 04:27:01 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071019SIM-VINYL.M
 Quant Title : SW846 8260
 QLast Update : Thu Jul 11 04:54:57 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.65	168	50790	5.00	ug/l	0.00
4) 1,4-Difluorobenzene	8.58	114	73137	5.00	ug/l	0.00
8) Chlorobenzene-d5	11.41	117	55125	5.00	ug/l	0.00
9) 1,4-Dichlorobenzene-d4	13.34	152	11198	5.00	ug/l	0.00

System Monitoring Compounds

3) 1,2-Dichloroethane-d4	8.01	65	29246	4.80	ug/l	0.00
Spiked Amount	5.000	Range	81 - 118	Recovery	=	96.00%
5) Dibromofluoromethane	7.59	113	22766	5.90	ug/l	0.00
Spiked Amount	5.000	Range	80 - 119	Recovery	=	118.00%
6) Toluene-d8	10.09	98	52287	5.60	ug/l	0.00
Spiked Amount	5.000	Range	89 - 112	Recovery	=	112.00%
7) 4-Bromofluorobenzene	12.39	95	16491	4.93	ug/l	0.00
Spiked Amount	5.000	Range	85 - 114	Recovery	=	98.60%

Target Compounds

	R.T.	QIon	Response	Conc	Units	Qvalue
2) Vinyl Chloride	2.17	62	25110	5.255	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN071119\
Data File : VN056659.D
Acq On : 11 Jul 2019 9:26
Operator : JC/SP
Sample : VSSTDCCC005
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VSSTDCCC005

Quant Time: Jul 12 04:27:01 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N071019SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Thu Jul 11 04:54:57 2019
Response via : Initial Calibration

