

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN071119\
Data File : VN056659.D
Acq On : 11 Jul 2019 9:26
Operator : JC/SP
Sample : VSSTDCCC005
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSSTDCCC005

Quant Time: Jul 12 04:27:01 2019
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N071019SIM-VINYL.M
Quant Title : SW846 8260
QLast Update : Thu Jul 11 04:54:57 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	99	0.00
2 C	Vinyl Chloride	0.470	0.494	-5.1#	103	0.00
3 S	1,2-Dichloroethane-d4	0.600	0.576	4.0	97	0.00
4 I	1,4-Difluorobenzene	1.000	1.000	0.0	82	0.00
5 S	Dibromofluoromethane	0.264	0.311	-17.8	100	0.00
6 S	Toluene-d8	0.638	0.715	-12.1	93	0.00
7 S	4-Bromofluorobenzene	0.229	0.225	1.7	86	0.00
8 I	Chlorobenzene-d5	1.000	1.000	0.0	94	0.00
9 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	78	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 1