

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071122\
 Data File : VN073409.D
 Acq On : 11 Jul 2022 10:44
 Operator : JC\MD
 Sample : VN0711MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0711MBL01

Quant Time: Jul 12 07:23:26 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.010	168	279918	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.898	114	460444	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	415996	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	197759	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	179138	44.331	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	88.660%
35) Dibromofluoromethane	7.951	113	151731	49.530	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	99.060%
50) Toluene-d8	10.375	98	484439	44.525	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.040%
62) 4-Bromofluorobenzene	12.669	95	207606	49.652	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	99.300%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071122\
Data File : VN073409.D
Acq On : 11 Jul 2022 10:44
Operator : JC\MD
Sample : VN0711MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0711MBL01

Quant Time: Jul 12 07:23:26 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

