

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071122\
 Data File : VN073410.D
 Acq On : 11 Jul 2022 11:08
 Operator : JC\MD
 Sample : VN0711WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0711WBL01

Quant Time: Jul 12 07:23:45 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	313043	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	520355	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.680	117	472075	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	222226	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	216735	47.960	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	95.920%
35) Dibromofluoromethane	7.951	113	178589	51.586	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.180%
50) Toluene-d8	10.380	98	586030	47.661	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.320%
62) 4-Bromofluorobenzene	12.669	95	244642	51.773	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	103.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071122\
Data File : VN073410.D
Acq On : 11 Jul 2022 11:08
Operator : JC\MD
Sample : VN0711WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0711WBL01

Quant Time: Jul 12 07:23:45 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

