

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071122\
 Data File : VN073420.D
 Acq On : 11 Jul 2022 15:30
 Operator : JC\MD
 Sample : N3651-12
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SW-4

Manual Integrations
 APPROVED

Reviewed By :Krupa Patel 07/12/2022
 Supervised By :Mahesh Dadoda 07/12/2022

Quant Time: Jul 12 07:27:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	271155	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.904	114	456444	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	424840	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	203050	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	179400	45.831	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery =	91.660%		
35) Dibromofluoromethane	7.951	113	148568	48.923	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	97.840%		
50) Toluene-d8	10.381	98	489932	45.424	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery =	90.840%		
62) 4-Bromofluorobenzene	12.669	95	210454	50.774	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	101.540%		
Target Compounds						
						Qvalue
11) Tert butyl alcohol	5.281	59	10428m	10.997	ug/l	
16) Acetone	4.228	43	6490m	2.954	ug/l	
19) Methyl tert-butyl Ether	5.546	73	11881	1.021	ug/l	# 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071122\
Data File : VN073420.D
Acq On : 11 Jul 2022 15:30
Operator : JC\MD
Sample : N3651-12
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SW-4

Quant Time: Jul 12 07:27:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Krupa Patel 07/12/2022
Supervised By :Mahesh Dadoda 07/12/2022

