

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071122\
 Data File : VN073421.D
 Acq On : 11 Jul 2022 15:55
 Operator : JC\MD
 Sample : N3651-13
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SW-5

Quant Time: Jul 12 07:28:04 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	260283	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	440898	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	391741	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.610	152	178524	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	175349	46.667	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.340%
35) Dibromofluoromethane	7.951	113	147040	50.127	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	100.260%
50) Toluene-d8	10.381	98	463534	44.492	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	88.980%
62) 4-Bromofluorobenzene	12.669	95	186515	46.586	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	93.180%
Target Compounds						
					Qvalue	
16) Acetone	4.234	43	10346	4.905	ug/l #	78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071122\
Data File : VN073421.D
Acq On : 11 Jul 2022 15:55
Operator : JC\MD
Sample : N3651-13
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SW-5

Quant Time: Jul 12 07:28:04 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

