

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071122\
 Data File : VN073422.D
 Acq On : 11 Jul 2022 16:19
 Operator : JC\MD
 Sample : N3651-14
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 SW-6

Quant Time: Jul 12 07:28:19 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Quant Title : SW846 8260
 QLast Update : Wed Jun 29 01:54:03 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.016	168	254981	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.904	114	430112	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.686	117	394378	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.615	152	185105	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.369	65	171843	46.685	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	93.360%
35) Dibromofluoromethane	7.951	113	140320	49.036	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	98.080%
50) Toluene-d8	10.380	98	453529	44.624	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	89.240%
62) 4-Bromofluorobenzene	12.668	95	192927	49.395	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	=	98.800%
Target Compounds						
16) Acetone	4.234	43	9434	4.566	ug/l	Qvalue # 78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071122\
Data File : VN073422.D
Acq On : 11 Jul 2022 16:19
Operator : JC\MD
Sample : N3651-14
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
SW-6

Quant Time: Jul 12 07:28:19 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
Quant Title : SW846 8260
QLast Update : Wed Jun 29 01:54:03 2022
Response via : Initial Calibration

