



Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071122\
 Data File : VN073407.D
 Acq On : 11 Jul 2022 09:25
 Operator : JC\MD
 Sample : BFB
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N062822W.M
 Title : SW846 8260
 Last Update : Wed Jun 29 01:54:03 2022



Spectrum Information: Average of 12.663 to 12.675 min.

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	21.2	14017	PASS
75	95	30	60	54.7	36085	PASS
95	95	100	100	100.0	65986	PASS
96	95	5	9	6.9	4556	PASS
173	174	0.00	2	1.5	759	PASS
174	95	50	100	77.7	51288	PASS
175	174	5	9	7.1	3635	PASS
176	174	95	101	97.0	49731	PASS
177	176	5	9	6.5	3240	PASS