

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071124\
 Data File : VN082865.D
 Acq On : 11 Jul 2024 10:32
 Operator : JC\MD
 Sample : VN0711MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0711MBL01

Quant Time: Jul 12 01:02:16 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	115876	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	207515	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	205599	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	89010	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	87519	53.880	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 107.760%		
35) Dibromofluoromethane	8.171	113	73000	54.822	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 109.640%		
50) Toluene-d8	10.571	98	262570	53.388	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 106.780%		
62) 4-Bromofluorobenzene	12.847	95	95958	54.317	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	= 108.640%		

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071124\
Data File : VN082865.D
Acq On : 11 Jul 2024 10:32
Operator : JC\MD
Sample : VN0711MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0711MBL01

Quant Time: Jul 12 01:02:16 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

