

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071124\
 Data File : VN082870.D
 Acq On : 11 Jul 2024 12:42
 Operator : JC\MD
 Sample : P3206-02 50X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 626-A

Quant Time: Jul 12 01:05:17 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	128050	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	234390	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	228600	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	99753	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	89324	49.763	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	99.520%
35) Dibromofluoromethane	8.165	113	73604	48.938	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	97.880%
50) Toluene-d8	10.571	98	265615	47.815	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	95.620%
62) 4-Bromofluorobenzene	12.853	95	96421	48.321	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	=	96.640%
Target Compounds						
16) Acetone	4.447	43	2734	4.326	ug/l	Qvalue # 59

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071124\
Data File : VN082870.D
Acq On : 11 Jul 2024 12:42
Operator : JC\MD
Sample : P3206-02 50X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
626-A

Quant Time: Jul 12 01:05:17 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

