

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071124\
 Data File : VN082872.D
 Acq On : 11 Jul 2024 13:31
 Operator : JC\MD
 Sample : P3187-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 WC-1

Quant Time: Jul 12 01:06:05 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	126796	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	237804	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	231483	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	107677	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	90986	51.190	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 102.380%		
35) Dibromofluoromethane	8.165	113	75692	49.604	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 99.200%		
50) Toluene-d8	10.571	98	284309	50.445	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 100.900%		
62) 4-Bromofluorobenzene	12.853	95	104701	51.717	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	= 103.440%		
Target Compounds						
						Qvalue
16) Acetone	4.430	43	51393	82.126	ug/l	97
20) Methylene Chloride	5.283	84	7201	4.765	ug/l #	71
25) 2-Butanone	7.482	43	5361	5.674	ug/l #	85
43) Isopropyl Acetate	8.688	43	145194	35.064	ug/l #	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071124\
Data File : VN082872.D
Acq On : 11 Jul 2024 13:31
Operator : JC\MD
Sample : P3187-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
WC-1

Quant Time: Jul 12 01:06:05 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

