

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071124\
 Data File : VN082873.D
 Acq On : 11 Jul 2024 13:55
 Operator : JC\MD
 Sample : P3183-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 N28660

Quant Time: Jul 12 01:06:31 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
 Quant Title : SW846 8260
 QLast Update : Mon Jul 08 13:46:04 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.230	168	124097	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.106	114	233320	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	228715	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	101407	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	88108	50.649	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 101.300%		
35) Dibromofluoromethane	8.165	113	71726	47.908	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 95.820%		
50) Toluene-d8	10.571	98	270412	48.902	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 97.800%		
62) 4-Bromofluorobenzene	12.847	95	99326	50.005	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	= 100.000%		
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	60725	99.149	ug/l	93
20) Methylene Chloride	5.277	84	6031	4.077	ug/l #	80
25) 2-Butanone	7.489	43	9372	10.134	ug/l	98
40) Benzene	8.612	78	17771	2.799	ug/l	99
43) Isopropyl Acetate	8.689	43	45492	9.050	ug/l #	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN071124\
Data File : VN082873.D
Acq On : 11 Jul 2024 13:55
Operator : JC\MD
Sample : P3183-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
N28660

Quant Time: Jul 12 01:06:31 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N070824W.M
Quant Title : SW846 8260
QLast Update : Mon Jul 08 13:46:04 2024
Response via : Initial Calibration

