

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN071124\  
 Data File : VN082883.D  
 Acq On : 11 Jul 2024 17:57  
 Operator : JC\MD  
 Sample : P3222-02  
 Misc : 5.0mL/MSVOA\_N/WATER  
 ALS Vial : 22 Sample Multiplier: 1

Instrument :  
 MSVOA\_N  
 ClientSampleId :  
 BP-VPB-196-EB-20240709

Quant Time: Jul 12 01:10:23 2024  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N070824W.M  
 Quant Title : SW846 8260  
 QLast Update : Mon Jul 08 13:46:04 2024  
 Response via : Initial Calibration

| Compound                    | R.T.           | QIon | Response            | Conc   | Units | Dev(Min)  |
|-----------------------------|----------------|------|---------------------|--------|-------|-----------|
| -----                       |                |      |                     |        |       |           |
| Internal Standards          |                |      |                     |        |       |           |
| 1) Pentafluorobenzene       | 8.224          | 168  | 115989              | 50.000 | ug/l  | 0.00      |
| 34) 1,4-Difluorobenzene     | 9.100          | 114  | 220464              | 50.000 | ug/l  | 0.00      |
| 63) Chlorobenzene-d5        | 11.865         | 117  | 223403              | 50.000 | ug/l  | 0.00      |
| 72) 1,4-Dichlorobenzene-d4  | 13.794         | 152  | 95298               | 50.000 | ug/l  | 0.00      |
|                             |                |      |                     |        |       |           |
| System Monitoring Compounds |                |      |                     |        |       |           |
| 33) 1,2-Dichloroethane-d4   | 8.583          | 65   | 86210               | 53.022 | ug/l  | 0.00      |
| Spiked Amount 50.000        | Range 74 - 125 |      | Recovery = 106.040% |        |       |           |
| 35) Dibromofluoromethane    | 8.171          | 113  | 72532               | 51.272 | ug/l  | 0.00      |
| Spiked Amount 50.000        | Range 75 - 124 |      | Recovery = 102.540% |        |       |           |
| 50) Toluene-d8              | 10.571         | 98   | 259131              | 49.594 | ug/l  | 0.00      |
| Spiked Amount 50.000        | Range 86 - 113 |      | Recovery = 99.180%  |        |       |           |
| 62) 4-Bromofluorobenzene    | 12.853         | 95   | 94436               | 50.316 | ug/l  | 0.00      |
| Spiked Amount 50.000        | Range 64 - 133 |      | Recovery = 100.640% |        |       |           |
|                             |                |      |                     |        |       |           |
| Target Compounds            |                |      |                     |        |       |           |
| 16) Acetone                 | 4.447          | 43   | 9265                | 16.185 | ug/l  | Qvalue 96 |
| -----                       |                |      |                     |        |       |           |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_N\Data\VN071124\  
Data File : VN082883.D  
Acq On : 11 Jul 2024 17:57  
Operator : JC\MD  
Sample : P3222-02  
Misc : 5.0mL/MSVOA\_N/WATER  
ALS Vial : 22 Sample Multiplier: 1

Instrument :  
MSVOA\_N  
ClientSampleId :  
BP-VPB-196-EB-20240709

Quant Time: Jul 12 01:10:23 2024  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_N\methods\82N070824W.M  
Quant Title : SW846 8260  
QLast Update : Mon Jul 08 13:46:04 2024  
Response via : Initial Calibration

